

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121223\
 Data File : PQ064452.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2023 13:15
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 23:29:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 09 05:18:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	2.773	179.4E6	209.0E6	43.253	49.778
2)	SA Decachlor...	8.696	7.570	142.4E6	236.4E6	48.581	58.573
Target Compounds							
16)	L4 AR-1242-1	4.576	3.783	46231148	49020356	446.077	472.406
17)	L4 AR-1242-2	4.595	3.799	68290380	72738432	440.190	470.826
18)	L4 AR-1242-3	4.654	3.960	43344115	39436990	447.499	480.879
19)	L4 AR-1242-4	4.746	4.046	34727425	37546619	456.812	481.200
20)	L4 AR-1242-5	5.469	4.544	33045451	46884190	398.066	486.178



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121223\
Data File : PQ064452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2023 13:15
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 23:29:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120823.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 09 05:18:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.460 min
Delta R.T.: 0.003 min
Response: 179439728
Conc: 43.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: 0.000 min
Response: 208989553
Conc: 49.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.696 min
Delta R.T.: 0.004 min
Response: 142412179
Conc: 48.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.570 min
Delta R.T.: 0.000 min
Response: 236412631
Conc: 58.57 ng/ml



#16 AR-1242-1


R.T.: 4.576 min
Delta R.T.: 0.004 min
Response: 46231148
Conc: 446.08 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 3.783 min
Delta R.T.: 0.000 min
Response: 49020356
Conc: 472.41 ng/ml




#17 AR-1242-2

R.T.: 4.595 min
Delta R.T.: 0.004 min
Response: 68290380
Conc: 440.19 ng/ml



#17 AR-1242-2


R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 72738432
Conc: 470.83 ng/ml



#18 AR-1242-3


R.T.: 4.654 min
Delta R.T.: 0.004 min
Response: 43344115
Conc: 447.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 3.960 min
Delta R.T.: 0.000 min
Response: 39436990
Conc: 480.88 ng/ml




#19 AR-1242-4

R.T.: 4.746 min
Delta R.T.: 0.003 min
Response: 34727425
Conc: 456.81 ng/ml



#19 AR-1242-4


R.T.: 4.046 min
Delta R.T.: 0.000 min
Response: 37546619
Conc: 481.20 ng/ml



#20 AR-1242-5

R.T.: 5.469 min
Delta R.T.: 0.004 min
Response: 33045451
Conc: 398.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 4.544 min
Delta R.T.: 0.000 min
Response: 46884190
Conc: 486.18 ng/ml