

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121224\
 Data File : PQ069648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2024 23:02
 Operator : YP\AJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 03:51:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 13 03:48:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.416	2.790	354.5E6	483.8E6	72.949	70.603
2) SA Decachlor...	8.527	7.567	371.3E6	745.0E6	143.063	139.513
Target Compounds						
26) L6 AR-1254-1	5.313	4.553	246.9E6	499.9E6	1378.046	1364.957
27) L6 AR-1254-2	5.528	4.700	385.4E6	435.1E6	1393.386	1352.840
28) L6 AR-1254-3	5.876	5.082	400.3E6	719.6E6	1415.350	1385.574
29) L6 AR-1254-4	6.157	5.308	279.8E6	429.6E6	1403.570	1370.902
30) L6 AR-1254-5	6.566	5.714	316.7E6	649.8E6	1415.146	1391.495



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121224\
Data File : PQ069648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2024 23:02
Operator : YP\AJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 03:51:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 13 03:48:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.416 min
Delta R.T.: 0.000 min
Response: 354521361
Conc: 72.95 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.790 min
Delta R.T.: 0.000 min
Response: 483787410
Conc: 70.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.527 min
Delta R.T.: 0.000 min
Response: 371342591
Conc: 143.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.567 min
Delta R.T.: 0.000 min
Response: 745021533
Conc: 139.51 ng/ml



#26 AR-1254-1


R.T.: 5.313 min
Delta R.T.: 0.000 min
Response: 246900765
Conc: 1378.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#26 AR-1254-1

R.T.: 4.553 min
Delta R.T.: 0.000 min
Response: 499936187
Conc: 1364.96 ng/ml




#27 AR-1254-2

R.T.: 5.528 min
Delta R.T.: 0.000 min
Response: 385414958
Conc: 1393.39 ng/ml



#27 AR-1254-2


R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 435134974
Conc: 1352.84 ng/ml



#28 AR-1254-3


R.T.: 5.876 min
Delta R.T.: 0.000 min
Response: 400274311
Conc: 1415.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#28 AR-1254-3

R.T.: 5.082 min
Delta R.T.: 0.000 min
Response: 719563324
Conc: 1385.57 ng/ml




#29 AR-1254-4

R.T.: 6.157 min
Delta R.T.: 0.000 min
Response: 279767626
Conc: 1403.57 ng/ml



#29 AR-1254-4


R.T.: 5.308 min
Delta R.T.: 0.000 min
Response: 429631081
Conc: 1370.90 ng/ml



#30 AR-1254-5

R.T.: 6.566 min
Delta R.T.: 0.001 min
Response: 316729053
Conc: 1415.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#30 AR-1254-5

R.T.: 5.714 min
Delta R.T.: 0.000 min
Response: 649844404
Conc: 1391.49 ng/ml