

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121318\
 Data File : PQ035455.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2018 17:35
 Operator : SM\SJ
 Sample : J6341-03 (Sig #1); J6314-03 (Sig #2)
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 121018-JETT-F-D-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 00:50:40 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 12 22:07:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.744	52159554	27644820	13.331	12.372
2) SA Decachlor...	10.148	8.732	46955817	27979707	15.930	15.830

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121318\
Data File : PQ035455.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2018 17:35
Operator : SM\SJ
Sample : J6341-03 (Sig #1); J6314-03 (Sig #2)
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
121018-JETT-F-D-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 00:50:40 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 12 22:07:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.436 min
Delta R.T.: -0.001 min
Response: 52159554
Conc: 13.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
121018-JETT-F-D-B




#1 Tetrachloro-m-xylene

R.T.: 3.744 min
Delta R.T.: 0.000 min
Response: 27644820
Conc: 12.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.148 min
Delta R.T.: -0.007 min
Response: 46955817
Conc: 15.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 27979707
Conc: 15.83 ng/ml