

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121318\
 Data File : PQ035456.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2018 17:51
 Operator : SM\SJ
 Sample : J6341-04 (Sig #1); J6314-04 (Sig #2)
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 121018-JETT-F-D-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 00:50:54 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 12 22:07:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.744	56679536	30297273	14.486	13.559
2) SA Decachlor...	10.149	8.733	40624032	24930258	13.782	14.105

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121318\
Data File : PQ035456.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2018 17:51
Operator : SM\SJ
Sample : J6341-04 (Sig #1); J6314-04 (Sig #2)
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
121018-JETT-F-D-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 00:50:54 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 12 22:07:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.436 min
Delta R.T.: 0.000 min
Response: 56679536
Conc: 14.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
121018-JETT-F-D-S




#1 Tetrachloro-m-xylene

R.T.: 3.744 min
Delta R.T.: 0.000 min
Response: 30297273
Conc: 13.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.149 min
Delta R.T.: -0.006 min
Response: 40624032
Conc: 13.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 24930258
Conc: 14.11 ng/ml