

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121323\
 Data File : PQ064500.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2023 23:15
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 23:36:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 09 05:18:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.456	2.773	185.0E6	236.5E6	44.590	56.325 #
2) SA Decachlor...	8.689	7.566	144.6E6	216.1E6	49.326	53.545
Target Compounds						
26) L6 AR-1254-1	5.401	4.542	72651764	112.8E6	474.327	538.905
27) L6 AR-1254-2	5.620	4.690	109.9E6	98933654	470.746	534.830
28) L6 AR-1254-3	5.976	5.073	114.5E6	147.7E6	465.732	521.969
29) L6 AR-1254-4	6.262	5.300	81884452	93881087	471.274	505.699
30) L6 AR-1254-5	6.677	5.708	91702062	135.1E6	475.134	520.427



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121323\
Data File : PQ064500.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2023 23:15
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 23:36:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120823.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 09 05:18:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.456 min
Delta R.T.: 0.000 min
Response: 184987059
Conc: 44.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.773 min
Delta R.T.: 0.000 min
Response: 236475896
Conc: 56.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.689 min
Delta R.T.: -0.003 min
Response: 144597290
Conc: 49.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.566 min
Delta R.T.: -0.004 min
Response: 216118758
Conc: 53.54 ng/ml



#26 AR-1254-1


R.T.: 5.401 min
Delta R.T.: 0.000 min
Response: 72651764
Conc: 474.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.542 min
Delta R.T.: -0.001 min
Response: 112838356
Conc: 538.91 ng/ml




#27 AR-1254-2

R.T.: 5.620 min
Delta R.T.: 0.001 min
Response: 109923620
Conc: 470.75 ng/ml



#27 AR-1254-2


R.T.: 4.690 min
Delta R.T.: -0.001 min
Response: 98933654
Conc: 534.83 ng/ml



#28 AR-1254-3


R.T.: 5.976 min
Delta R.T.: 0.000 min
Response: 114474488
Conc: 465.73 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.073 min
Delta R.T.: -0.001 min
Response: 147670254
Conc: 521.97 ng/ml




#29 AR-1254-4

R.T.: 6.262 min
Delta R.T.: 0.000 min
Response: 81884452
Conc: 471.27 ng/ml



#29 AR-1254-4


R.T.: 5.300 min
Delta R.T.: -0.002 min
Response: 93881087
Conc: 505.70 ng/ml



#30 AR-1254-5

R.T.: 6.677 min
Delta R.T.: 0.000 min
Response: 91702062
Conc: 475.13 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 135129170
Conc: 520.43 ng/ml