

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121523\
 Data File : PQ064551.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2023 11:53
 Operator : YP\AJ
 Sample : 05557-18
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 600X-6C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 13:37:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 09 05:18:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.456	2.773	67313212	78679524	16.225	18.740
2) SA Decachlor...	8.691	7.565	40120450	45723031	13.686	11.328
Target Compounds						
26) L6 AR-1254-1	5.401	4.542	942.9E6	1373.0E6	6156.064	6557.308
27) L6 AR-1254-2	5.619	4.690	2014.7E6	1601.3E6	8627.776	8656.669
28) L6 AR-1254-3	5.975	5.072	3175.6E6	3420.4E6	12919.900	12090.139
29) L6 AR-1254-4	6.262	5.300	2726.7E6	2616.7E6	15693.241	14094.921
30) L6 AR-1254-5	6.678	5.707	4018.7E6	4983.6E6	20822.038	19193.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121523\
Data File : PQ064551.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 11:53
Operator : YP\AJ
Sample : 05557-18
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
600X-6C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 13:37:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ120823.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 09 05:18:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

