

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121918\
 Data File : PQ035790.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2018 02:43
 Operator : SM\SJ
 Sample : J6445-04
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 EB02_6-8

Manual Integrations
 APPROVED

Sohil
 12/20/2018 4:20:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 06:07:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 12 22:07:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.739	70309136	42050712	17.970	18.820
2) SA Decachlor...	10.114	8.713	42099988	24610593	14.283m	13.924

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121918\
Data File : PQ035790.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2018 02:43
Operator : SM\SJ
Sample : J6445-04
Misc :
ALS Vial : 55 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
EB02_6-8


Manual Integrations
APPROVED

Sohil
12/20/2018 4:20:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 06:07:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 12 22:07:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.424 min
Delta R.T.: -0.013 min
Response: 70309136
Conc: 17.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
EB02_6-8


Manual Integrations
APPROVED

Sohil
12/20/2018 4:20:39 PM




#1 Tetrachloro-m-xylene

R.T.: 3.739 min
Delta R.T.: -0.005 min
Response: 42050712
Conc: 18.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.114 min
Delta R.T.: -0.041 min
Response: 42099988
Conc: 14.28 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.713 min
Delta R.T.: -0.019 min
Response: 24610593
Conc: 13.92 ng/ml