

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ122120\
 Data File : PQ051424.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2020 23:52
 Operator : DD\AJ
 Sample : AR1254ICC400
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254301

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 04:51:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ122120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 22 04:49:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.236	4.255	576.9E6	163.3E6	20.000	20.000
2) SA Decachlor...	11.431	9.607	785.6E6	257.4E6	40.000	40.000
Target Compounds						
26) L6 AR-1254-1	7.469	6.389	399.0E6	171.1E6	400.000	400.000
27) L6 AR-1254-2	7.697	6.547	616.6E6	147.6E6	400.000	400.000
28) L6 AR-1254-3	8.079	6.969	668.9E6	245.9E6	400.000	400.000
29) L6 AR-1254-4	8.375	7.206	481.7E6	152.2E6	400.000	400.000
30) L6 AR-1254-5	8.804	7.637	511.2E6	212.1E6	400.000	400.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ122120\
Data File : PQ051424.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2020 23:52
Operator : DD\AJ
Sample : AR1254ICC400
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254301

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 04:51:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ122120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 22 04:49:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

