

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ122421\
 Data File : PQ055353.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Dec 2021 12:45
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ122421AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 22:28:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 22:17:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.244	4.315	655.6E6	540.3E6	53.077	54.952
2)	SA Decachlor...	11.435	9.707	840.9E6	451.5E6	52.911	53.475
Target Compounds							
26)	L6 AR-1254-1	7.479	6.463	191.8E6	288.1E6	516.297	531.448
27)	L6 AR-1254-2	7.708	6.621	302.3E6	246.2E6	520.315	526.887
28)	L6 AR-1254-3	8.092	7.046	367.7E6	406.3E6	521.978	534.113
29)	L6 AR-1254-4	8.388	7.283	277.8E6	249.1E6	525.404	535.425
30)	L6 AR-1254-5	8.818	7.716	297.9E6	353.2E6	524.045	548.604

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ122421\
Data File : PQ055353.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Dec 2021 12:45
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ122421AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 22:28:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 22:17:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

