

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010219\
 Data File : PQ036200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jan 2019 16:45
 Operator : SM\SJ
 Sample : J6379-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P001-SS003-0612-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 02 17:21:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ122918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 31 04:01:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.401	3.727	85350394	48871392	16.678	17.392
2) SA Decachlor...	10.074	8.697	42038938	23586424	12.052	11.817
Target Compounds						
31) L7 AR-1260-1	7.159	6.272	217.5E6	144.8E6	1092.025	1082.484
32) L7 AR-1260-2	7.415	6.460	258.0E6	171.7E6	1093.415	1078.809
33) L7 AR-1260-3	7.774	6.612	195.0E6	147.1E6	1415.849	1002.591 #
34) L7 AR-1260-4	8.002	7.080	228.1E6	136.2E6	1350.933	1408.127
35) L7 AR-1260-5	8.319	7.323	424.4E6	295.0E6	1308.032	1263.683

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010219\
Data File : PQ036200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jan 2019 16:45
Operator : SM\SJ
Sample : J6379-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
P001-SS003-0612-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 02 17:21:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ122918.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 31 04:01:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

