

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ010819.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 08 12:48:10 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ036268.D 1000 =PQ036266.D 750 =PQ036267.D
 250 =PQ036269.D 50 =PQ036270.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.764	4.523	5.042	4.519	4.010	4.572	E6	8.32
2) SA	Decachlorobiphenyl	3.622	3.608	3.559	3.621	3.777	3.637	E6	2.25
3) L1	AR-1016-1	1.404	1.306	1.485	1.400	1.401	1.399	E5	4.53
4) L1	AR-1016-2	2.112	2.009	2.313	2.109	2.131	2.135	E5	5.18
5) L1	AR-1016-3	1.231	1.161	1.323	1.229	1.259	1.241	E5	4.72
6) L1	AR-1016-4	1.013	0.959	1.100	1.008	0.983	1.013	E5	5.25
7) L1	AR-1016-5	0.991	0.927	1.076	1.013	0.986	0.999	E5	5.35
8) L2	AR-1221-1	3.534					3.534	E4	0.00
9) L2	AR-1221-2	2.656					2.656	E4	0.00
10) L2	AR-1221-3	8.915					8.915	E4	0.00
11) L3	AR-1232-1	7.618					7.618	E4	0.00
12) L3	AR-1232-2	3.817					3.817	E4	0.00
13) L3	AR-1232-3	8.047					8.047	E4	0.00
14) L3	AR-1232-4	3.844					3.844	E4	0.00
15) L3	AR-1232-5	2.864					2.864	E4	0.00
16) L4	AR-1242-1	9.942					9.942	E4	0.00
17) L4	AR-1242-2	1.511					1.511	E5	0.00
18) L4	AR-1242-3	8.869					8.869	E4	0.00
19) L4	AR-1242-4	7.285					7.285	E4	0.00
20) L4	AR-1242-5	7.935					7.935	E4	0.00
21) L5	AR-1248-1	1.001					1.001	E5	0.00
22) L5	AR-1248-2	1.117					1.117	E5	0.00
23) L5	AR-1248-3	1.253					1.253	E5	0.00
24) L5	AR-1248-4	1.426					1.426	E5	0.00
25) L5	AR-1248-5	1.367					1.367	E5	0.00
26) L6	AR-1254-1	1.359					1.359	E5	0.00
27) L6	AR-1254-2	2.127					2.127	E5	0.00
28) L6	AR-1254-3	2.347					2.347	E5	0.00
29) L6	AR-1254-4	1.769					1.769	E5	0.00
30) L6	AR-1254-5	2.048					2.048	E5	0.00
31) L7	AR-1260-1	1.848	1.825	2.045	1.877	1.924	1.904	E5	4.58
32) L7	AR-1260-2	2.220	2.216	2.429	2.247	2.363	2.295	E5	4.18
33) L7	AR-1260-3	1.358	1.355	1.462	1.389	1.476	1.408	E5	4.08
34) L7	AR-1260-4	1.672	1.691	1.768	1.692	1.799	1.724	E5	3.22
35) L7	AR-1260-5	3.292	3.413	3.430	3.260	3.387	3.357	E5	2.26
36) L8	AR-1262-1	2.343					2.343	E5	0.00
37) L8	AR-1262-2	4.291					4.291	E5	0.00
38) L8	AR-1262-3	9.306					9.306	E4	0.00
39) L8	AR-1262-4	2.074					2.074	E5	0.00
40) L8	AR-1262-5	1.453					1.453	E5	0.00
41) L9	AR-1268-1	5.079					5.079	E5	0.00
42) L9	AR-1268-2	4.782					4.782	E5	0.00
43) L9	AR-1268-3	4.096					4.096	E5	0.00
44) L9	AR-1268-4	1.642					1.642	E5	0.00
45) L9	AR-1268-5	1.356					1.356	E6	0.00

Signal #2 Calibration Files

500 =PQ036268.D 1000 =PQ036266.D 750 =PQ036267.D
 250 =PQ036269.D 50 =PQ036270.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.522	2.344	2.631	2.415	2.191	2.420	E6	6.95
2) SA	Decachlorobiphenyl	1.875	1.814	1.794	1.909	2.220	1.922	E6	8.99
3) L1	AR-1016-1	7.873	7.097	8.040	7.878	7.919	7.761	E4	4.86
4) L1	AR-1016-2	1.145	1.059	1.236	1.159	1.095	1.139	E5	5.93
5) L1	AR-1016-3	5.866	5.320	6.181	5.930	5.629	5.785	E4	5.64

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ010819.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 08 12:48:10 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ036268.D 1000 =PQ036266.D 750 =PQ036267.D
 250 =PQ036269.D 50 =PQ036270.D

Compound			500	1000	750	250	50	Avg		%RSD
6)	L1	AR-1016-4	4.746	4.232	4.976	4.974	4.641	4.714	E4	6.50
7)	L1	AR-1016-5	6.079	5.482	6.374	6.211	6.271	6.084	E4	5.80
8)	L2	AR-1221-1	1.962					1.962	E4	0.00
9)	L2	AR-1221-2	1.482					1.482	E4	0.00
10)	L2	AR-1221-3	4.944					4.944	E4	0.00
11)	L3	AR-1232-1	4.303					4.303	E4	0.00
12)	L3	AR-1232-2	4.505					4.505	E4	0.00
13)	L3	AR-1232-3	2.266					2.266	E4	0.00
14)	L3	AR-1232-4	2.058					2.058	E4	0.00
15)	L3	AR-1232-5	2.193					2.193	E4	0.00
16)	L4	AR-1242-1	5.765					5.765	E4	0.00
17)	L4	AR-1242-2	8.155					8.155	E4	0.00
18)	L4	AR-1242-3	4.275					4.275	E4	0.00
19)	L4	AR-1242-4	4.255					4.255	E4	0.00
20)	L4	AR-1242-5	5.432					5.432	E4	0.00
21)	L5	AR-1248-1	4.693					4.693	E4	0.00
22)	L5	AR-1248-2	6.276					6.276	E4	0.00
23)	L5	AR-1248-3	6.425					6.425	E4	0.00
24)	L5	AR-1248-4	7.681					7.681	E4	0.00
25)	L5	AR-1248-5	7.632					7.632	E4	0.00
26)	L6	AR-1254-1	1.046					1.046	E5	0.00
27)	L6	AR-1254-2	9.138					9.138	E4	0.00
28)	L6	AR-1254-3	1.592					1.592	E5	0.00
29)	L6	AR-1254-4	1.050					1.050	E5	0.00
30)	L6	AR-1254-5	1.453					1.453	E5	0.00
31)	L7	AR-1260-1	1.147	1.084	1.215	1.176	1.332	1.191	E5	7.73
32)	L7	AR-1260-2	1.380	1.326	1.448	1.416	1.608	1.436	E5	7.40
33)	L7	AR-1260-3	1.284	1.247	1.355	1.305	1.410	1.320	E5	4.82
34)	L7	AR-1260-4	8.755	8.461	8.847	8.901	9.982	8.989	E4	6.46
35)	L7	AR-1260-5	2.206	2.171	2.199	2.150	2.320	2.209	E5	2.97
36)	L8	AR-1262-1	1.542					1.542	E5	0.00
37)	L8	AR-1262-2	2.784					2.784	E5	0.00
38)	L8	AR-1262-3	1.037					1.037	E5	0.00
39)	L8	AR-1262-4	1.990					1.990	E5	0.00
40)	L8	AR-1262-5	8.793					8.793	E4	0.00
41)	L9	AR-1268-1	3.177					3.177	E5	0.00
42)	L9	AR-1268-2	2.880					2.880	E5	0.00
43)	L9	AR-1268-3	2.354					2.354	E5	0.00
44)	L9	AR-1268-4	9.943					9.943	E4	0.00
45)	L9	AR-1268-5	7.516					7.516	E5	0.00

(#) = Out of Range