

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010819\
 Data File : PQ036299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2019 02:04
 Operator : SM\SJ
 Sample : K1072-04
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SWITCH-PIERS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 03:37:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ010819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 09 01:06:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.388	3.723	75824203	46064069	16.586	19.031
2) SA Decachlor...	10.036	8.679	60388159	30913834	16.602	16.081
Target Compounds						
34) L7 AR-1260-4	0.000	7.093	0	2225635	N.D.	24.759 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010819\
Data File : PQ036299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2019 02:04
Operator : SM\SJ
Sample : K1072-04
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SWITCH-PIERS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 03:37:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ010819.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 09 01:06:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.388 min
Delta R.T.: -0.002 min
Response: 75824203
Conc: 16.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SWITCH-PIERS




#1 Tetrachloro-m-xylene

R.T.: 3.723 min
Delta R.T.: 0.002 min
Response: 46064069
Conc: 19.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.036 min
Delta R.T.: -0.009 min
Response: 60388159
Conc: 16.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.004 min
Response: 30913834
Conc: 16.08 ng/ml



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.984 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
SWITCH-PIERS



#34 AR-1260-4

R.T.: 7.093 min
Delta R.T.: 0.023 min
Response: 2225635
Conc: 24.76 ng/ml