

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010819\
 Data File : PQ036300.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2019 02:20
 Operator : SM\SJ
 Sample : K1072-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 26KV-POTHEAD-SLABS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 03:37:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ010819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 09 01:06:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.388	3.724	85970456	52102300	18.806	21.526
2) SA Decachlor...	10.035	8.680	67978771	34346023	18.688	17.866

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010819\
Data File : PQ036300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2019 02:20
Operator : SM\SJ
Sample : K1072-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
26KV-POTHEAD-SLABS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 03:37:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ010819.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 09 01:06:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.388 min
Delta R.T.: -0.002 min
Response: 85970456
Conc: 18.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
26KV-POTHEAD-SLABS




#1 Tetrachloro-m-xylene

R.T.: 3.724 min
Delta R.T.: 0.002 min
Response: 52102300
Conc: 21.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.035 min
Delta R.T.: -0.010 min
Response: 67978771
Conc: 18.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.680 min
Delta R.T.: -0.004 min
Response: 34346023
Conc: 17.87 ng/ml