

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020419\
 Data File : PQ036967.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2019 16:56
 Operator : SM\SJ
 Sample : PB116791BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PB116791BL

Manual Integrations
 APPROVED

Sohil
 2/5/2019 12:08:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 02:20:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 24 06:01:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.327	3.696	106.1E6	53353348	23.045	22.268m
2) SA Decachlor...	9.907	8.630	76993857	36833584	21.119m	19.579m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020419\
Data File : PQ036967.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2019 16:56
Operator : SM\SJ
Sample : PB116791BL
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
PB116791BL


Manual Integrations
APPROVED

Sohil
2/5/2019 12:08:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 02:20:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012419.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 24 06:01:26 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.327 min
Delta R.T.: -0.026 min
Response: 106072610
Conc: 23.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB116791BL


Manual Integrations
APPROVED

Sohil
2/5/2019 12:08:34 PM




#1 Tetrachloro-m-xylene

R.T.: 3.696 min
Delta R.T.: -0.010 min
Response: 53353348
Conc: 22.27 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.907 min
Delta R.T.: -0.057 min
Response: 76993857
Conc: 21.12 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.630 min
Delta R.T.: -0.022 min
Response: 36833584
Conc: 19.58 ng/ml m