

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ020619.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 06 18:32:04 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ036991.D 1000 =PQ036989.D 750 =PQ036990.D
 250 =PQ036992.D 50 =PQ036993.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.434	4.809	4.408	4.149	4.415	4.443	E6	5.31
2) SA	Decachlorobiphenyl	4.091	3.894	4.000	4.039	4.591	4.123	E6	6.58
3) L1	AR-1016-1	1.350	1.425	1.296	1.308	1.506	1.377	E5	6.39
4) L1	AR-1016-2	2.003	2.115	2.003	1.971	2.365	2.091	E5	7.77
5) L1	AR-1016-3	1.161	1.211	1.138	1.130	1.342	1.196	E5	7.29
6) L1	AR-1016-4	0.968	1.006	0.956	0.934	1.073	0.987	E5	5.51
7) L1	AR-1016-5	0.953	0.982	0.932	0.966	1.070	0.981	E5	5.43
8) L2	AR-1221-1	7.652					7.652	E4	0.00
9) L2	AR-1221-2	5.278					5.278	E4	0.00
10) L2	AR-1221-3	1.847					1.847	E5	0.00
11) L3	AR-1232-1	1.017					1.017	E5	0.00
12) L3	AR-1232-2	5.171					5.171	E4	0.00
13) L3	AR-1232-3	1.078					1.078	E5	0.00
14) L3	AR-1232-4	5.154					5.154	E4	0.00
15) L3	AR-1232-5	3.750					3.750	E4	0.00
16) L4	AR-1242-1	1.243					1.243	E5	0.00
17) L4	AR-1242-2	1.911					1.911	E5	0.00
18) L4	AR-1242-3	1.097					1.097	E5	0.00
19) L4	AR-1242-4	9.061					9.061	E4	0.00
20) L4	AR-1242-5	9.899					9.899	E4	0.00
21) L5	AR-1248-1	8.698					8.698	E4	0.00
22) L5	AR-1248-2	1.236					1.236	E5	0.00
23) L5	AR-1248-3	1.378					1.378	E5	0.00
24) L5	AR-1248-4	1.613					1.613	E5	0.00
25) L5	AR-1248-5	1.531					1.531	E5	0.00
26) L6	AR-1254-1	4.359					4.359	E5	0.00
27) L6	AR-1254-2	6.782					6.782	E5	0.00
28) L6	AR-1254-3	7.014					7.014	E5	0.00
29) L6	AR-1254-4	5.314					5.314	E5	0.00
30) L6	AR-1254-5	5.438					5.438	E5	0.00
31) L7	AR-1260-1	1.927	1.859	1.879	1.917	2.255	1.967	E5	8.29
32) L7	AR-1260-2	2.491	2.399	2.449	2.485	2.801	2.525	E5	6.28
33) L7	AR-1260-3	1.601	1.505	1.561	1.587	1.756	1.602	E5	5.85
34) L7	AR-1260-4	1.818	1.725	1.779	1.795	1.950	1.813	E5	4.62
35) L7	AR-1260-5	3.873	3.706	3.828	3.758	3.996	3.832	E5	2.92
36) L8	AR-1262-1	2.957					2.957	E5	0.00
37) L8	AR-1262-2	5.346					5.346	E5	0.00
38) L8	AR-1262-3	2.308					2.308	E5	0.00
39) L8	AR-1262-4	2.698					2.698	E5	0.00
40) L8	AR-1262-5	1.884					1.884	E5	0.00
41) L9	AR-1268-1	6.694					6.694	E5	0.00
42) L9	AR-1268-2	6.209					6.209	E5	0.00
43) L9	AR-1268-3	5.490					5.490	E5	0.00
44) L9	AR-1268-4	2.070					2.070	E5	0.00
45) L9	AR-1268-5	1.671					1.671	E6	0.00

Signal #2 Calibration Files

500 =PQ036991.D 1000 =PQ036989.D 750 =PQ036990.D
 250 =PQ036992.D 50 =PQ036993.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.303	2.453	2.257	2.169	2.365	2.309	E6	4.65
2) SA	Decachlorobiphenyl	1.839	1.666	1.761	1.870	2.260	1.879	E6	12.08
3) L1	AR-1016-1	7.065	7.096	6.785	7.102	8.088	7.227	E4	6.90
4) L1	AR-1016-2	1.075	1.105	1.047	1.051	1.137	1.083	E5	3.49
5) L1	AR-1016-3	5.400	5.384	5.204	5.424	6.129	5.508	E4	6.50

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ020619.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 06 18:32:04 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ036991.D 1000 =PQ036989.D 750 =PQ036990.D
 250 =PQ036992.D 50 =PQ036993.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	4.399	4.285	4.197	4.535	4.982	4.480	E4	6.88
7)	L1 AR-1016-5	5.641	5.514	5.406	5.725	6.748	5.807	E4	9.30
8)	L2 AR-1221-1	4.034					4.034	E4	0.00
9)	L2 AR-1221-2	2.861					2.861	E4	0.00
10)	L2 AR-1221-3	9.914					9.914	E4	0.00
11)	L3 AR-1232-1	5.522					5.522	E4	0.00
12)	L3 AR-1232-2	5.908					5.908	E4	0.00
13)	L3 AR-1232-3	2.961					2.961	E4	0.00
14)	L3 AR-1232-4	2.643					2.643	E4	0.00
15)	L3 AR-1232-5	2.860					2.860	E4	0.00
16)	L4 AR-1242-1	6.726					6.726	E4	0.00
17)	L4 AR-1242-2	1.025					1.025	E5	0.00
18)	L4 AR-1242-3	5.107					5.107	E4	0.00
19)	L4 AR-1242-4	4.979					4.979	E4	0.00
20)	L4 AR-1242-5	5.912					5.912	E4	0.00
21)	L5 AR-1248-1	4.937					4.937	E4	0.00
22)	L5 AR-1248-2	6.591					6.591	E4	0.00
23)	L5 AR-1248-3	6.666					6.666	E4	0.00
24)	L5 AR-1248-4	8.084					8.084	E4	0.00
25)	L5 AR-1248-5	7.906					7.906	E4	0.00
26)	L6 AR-1254-1	3.356					3.356	E5	0.00
27)	L6 AR-1254-2	2.805					2.805	E5	0.00
28)	L6 AR-1254-3	4.555					4.555	E5	0.00
29)	L6 AR-1254-4	2.831					2.831	E5	0.00
30)	L6 AR-1254-5	3.699					3.699	E5	0.00
31)	L7 AR-1260-1	1.088	1.008	1.055	1.116	1.309	1.115	E5	10.36
32)	L7 AR-1260-2	1.335	1.215	1.292	1.356	1.600	1.360	E5	10.66
33)	L7 AR-1260-3	1.233	1.131	1.203	1.245	1.407	1.244	E5	8.15
34)	L7 AR-1260-4	8.457	7.547	8.116	8.572	9.678	8.474	E4	9.23
35)	L7 AR-1260-5	2.073	1.876	2.013	2.050	2.325	2.068	E5	7.88
36)	L8 AR-1262-1	1.624					1.624	E5	0.00
37)	L8 AR-1262-2	2.881					2.881	E5	0.00
38)	L8 AR-1262-3	1.156					1.156	E5	0.00
39)	L8 AR-1262-4	2.092					2.092	E5	0.00
40)	L8 AR-1262-5	9.285					9.285	E4	0.00
41)	L9 AR-1268-1	3.399					3.399	E5	0.00
42)	L9 AR-1268-2	3.094					3.094	E5	0.00
43)	L9 AR-1268-3	2.558					2.558	E5	0.00
44)	L9 AR-1268-4	1.047					1.047	E5	0.00
45)	L9 AR-1268-5	8.074					8.074	E5	0.00

(#) = Out of Range