

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021119\
 Data File : PQ037180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Feb 2019 01:41
 Operator : SM\SJ
 Sample : K1452-01
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 TEST-PIT-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 03:25:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ020619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 06 18:32:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.289	3.682	121.7E6	63683001	27.385	27.575
2) SA Decachlor...	9.821	8.598	67088399	29763109	16.272m	15.841
Target Compounds						
26) L6 AR-1254-1	6.277	5.534	6222413	5205249	14.275m	15.512
27) L6 AR-1254-2	6.493	5.679	11615495	4218140	17.128m	15.040
28) L6 AR-1254-3	6.861	6.078	11330949	7908818	16.156m	17.362m
29) L6 AR-1254-4	7.143	6.304	7888431	4045282	14.846	14.289
30) L6 AR-1254-5	7.557	6.714	8450128	6004515	15.539m	16.231



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021119\
Data File : PQ037180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2019 01:41
Operator : SM\SJ
Sample : K1452-01
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
TEST-PIT-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 03:25:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ020619.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 06 18:32:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.289 min
Delta R.T.: -0.033 min
Response: 121677237
Conc: 27.39 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TEST-PIT-2




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.012 min
Response: 63683001
Conc: 27.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.821 min
Delta R.T.: -0.072 min
Response: 67088399
Conc: 16.27 ng/ml m



#2 Decachlorobiphenyl


R.T.: 8.598 min
Delta R.T.: -0.027 min
Response: 29763109
Conc: 15.84 ng/ml



#26 AR-1254-1


R.T.: 6.277 min
Delta R.T.: -0.040 min
Response: 6222413
Conc: 14.28 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
TEST-PIT-2




#26 AR-1254-1

R.T.: 5.534 min
Delta R.T.: -0.018 min
Response: 5205249
Conc: 15.51 ng/ml




#27 AR-1254-2

R.T.: 6.493 min
Delta R.T.: -0.038 min
Response: 11615495
Conc: 17.13 ng/ml m



#27 AR-1254-2


R.T.: 5.679 min
Delta R.T.: -0.017 min
Response: 4218140
Conc: 15.04 ng/ml



#28 AR-1254-3


R.T.: 6.861 min
Delta R.T.: -0.040 min
Response: 11330949
Conc: 16.16 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
TEST-PIT-2




#28 AR-1254-3

R.T.: 6.078 min
Delta R.T.: -0.018 min
Response: 7908818
Conc: 17.36 ng/ml m




#29 AR-1254-4

R.T.: 7.143 min
Delta R.T.: -0.040 min
Response: 7888431
Conc: 14.85 ng/ml



#29 AR-1254-4


R.T.: 6.304 min
Delta R.T.: -0.018 min
Response: 4045282
Conc: 14.29 ng/ml



#30 AR-1254-5

R.T.: 7.557 min
Delta R.T.: -0.042 min
Response: 8450128
Conc: 15.54 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
TEST-PIT-2



#30 AR-1254-5

R.T.: 6.714 min
Delta R.T.: -0.020 min
Response: 6004515
Conc: 16.23 ng/ml