

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ021319.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 13 16:51:51 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ037209.D 1000 =PQ037207.D 750 =PQ037208.D
 250 =PQ037210.D 50 =PQ037211.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	6.110	6.034	6.123	5.612	5.281	5.832	E6	6.38
2) SA	Decachlorobiphenyl	4.796	4.621	4.707	4.953	5.412	4.898	E6	6.38
3) L1	AR-1016-1	1.826	1.711	1.796	1.769	1.749	1.770	E5	2.48
4) L1	AR-1016-2	2.828	2.624	2.779	2.830	2.538	2.720	E5	4.86
5) L1	AR-1016-3	1.664	1.520	1.627	1.671	1.495	1.595	E5	5.17
6) L1	AR-1016-4	1.381	1.272	1.364	1.375	1.352	1.349	E5	3.27
7) L1	AR-1016-5	1.353	1.212	1.314	1.362	1.238	1.296	E5	5.23
8) L2	AR-1221-1	9.924					9.924	E4	0.00
9) L2	AR-1221-2	6.914					6.914	E4	0.00
10) L2	AR-1221-3	2.434					2.434	E5	0.00
11) L3	AR-1232-1	9.698					9.698	E4	0.00
12) L3	AR-1232-2	4.976					4.976	E4	0.00
13) L3	AR-1232-3	1.055					1.055	E5	0.00
14) L3	AR-1232-4	4.944					4.944	E4	0.00
15) L3	AR-1232-5	3.604					3.604	E4	0.00
16) L4	AR-1242-1	1.165					1.165	E5	0.00
17) L4	AR-1242-2	1.834					1.834	E5	0.00
18) L4	AR-1242-3	1.079					1.079	E5	0.00
19) L4	AR-1242-4	8.835					8.835	E4	0.00
20) L4	AR-1242-5	9.959					9.959	E4	0.00
21) L5	AR-1248-1	1.120					1.120	E5	0.00
22) L5	AR-1248-2	1.576					1.576	E5	0.00
23) L5	AR-1248-3	1.763					1.763	E5	0.00
24) L5	AR-1248-4	2.037					2.037	E5	0.00
25) L5	AR-1248-5	1.935					1.935	E5	0.00
26) L6	AR-1254-1	5.761					5.761	E5	0.00
27) L6	AR-1254-2	9.063					9.063	E5	0.00
28) L6	AR-1254-3	9.810					9.810	E5	0.00
29) L6	AR-1254-4	7.393					7.393	E5	0.00
30) L6	AR-1254-5	7.595					7.595	E5	0.00
31) L7	AR-1260-1	2.724	2.345	2.568	2.850	2.599	2.617	E5	7.22
32) L7	AR-1260-2	3.463	2.981	3.228	3.620	3.338	3.326	E5	7.28
33) L7	AR-1260-3	2.202	1.911	2.045	2.372	2.199	2.146	E5	8.16
34) L7	AR-1260-4	2.455	2.138	2.241	2.626	2.522	2.396	E5	8.42
35) L7	AR-1260-5	5.034	4.593	4.731	5.382	5.084	4.965	E5	6.26
36) L8	AR-1262-1	4.312					4.312	E5	0.00
37) L8	AR-1262-2	7.413					7.413	E5	0.00
38) L8	AR-1262-3	3.119					3.119	E5	0.00
39) L8	AR-1262-4	3.648					3.648	E5	0.00
40) L8	AR-1262-5	2.265					2.265	E5	0.00
41) L9	AR-1268-1	7.651					7.651	E5	0.00
42) L9	AR-1268-2	7.018					7.018	E5	0.00
43) L9	AR-1268-3	6.141					6.141	E5	0.00
44) L9	AR-1268-4	2.276					2.276	E5	0.00
45) L9	AR-1268-5	1.807					1.807	E6	0.00

Signal #2 Calibration Files

500 =PQ037209.D 1000 =PQ037207.D 750 =PQ037208.D
 250 =PQ037210.D 50 =PQ037211.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	3.144	3.101	3.176	2.917	2.814	3.030	E6	5.19
2) SA	Decachlorobiphenyl	2.119	2.031	2.094	2.184	2.438	2.173	E6	7.27
3) L1	AR-1016-1	9.587	8.855	9.582	9.402	9.587	9.403	E4	3.36
4) L1	AR-1016-2	1.464	1.371	1.424	1.444	1.329	1.406	E5	3.93
5) L1	AR-1016-3	7.356	6.742	7.159	7.233	6.653	7.029	E4	4.44

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ021319.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 13 16:51:51 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ037209.D 1000 =PQ037207.D 750 =PQ037208.D
 250 =PQ037210.D 50 =PQ037211.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	5.838	5.244	5.621	5.811	5.159	5.535	E4	5.73
7)	L1 AR-1016-5	7.528	6.770	7.307	7.445	7.060	7.222	E4	4.27
8)	L2 AR-1221-1	5.168					5.168	E4	0.00
9)	L2 AR-1221-2	3.620					3.620	E4	0.00
10)	L2 AR-1221-3	1.273					1.273	E5	0.00
11)	L3 AR-1232-1	5.129					5.129	E4	0.00
12)	L3 AR-1232-2	5.484					5.484	E4	0.00
13)	L3 AR-1232-3	2.676					2.676	E4	0.00
14)	L3 AR-1232-4	2.308					2.308	E4	0.00
15)	L3 AR-1232-5	2.494					2.494	E4	0.00
16)	L4 AR-1242-1	6.221					6.221	E4	0.00
17)	L4 AR-1242-2	9.618					9.618	E4	0.00
18)	L4 AR-1242-3	4.755					4.755	E4	0.00
19)	L4 AR-1242-4	4.512					4.512	E4	0.00
20)	L4 AR-1242-5	5.677					5.677	E4	0.00
21)	L5 AR-1248-1	5.951					5.951	E4	0.00
22)	L5 AR-1248-2	8.016					8.016	E4	0.00
23)	L5 AR-1248-3	8.106					8.106	E4	0.00
24)	L5 AR-1248-4	9.795					9.795	E4	0.00
25)	L5 AR-1248-5	9.644					9.644	E4	0.00
26)	L6 AR-1254-1	4.230					4.230	E5	0.00
27)	L6 AR-1254-2	3.574					3.574	E5	0.00
28)	L6 AR-1254-3	5.898					5.898	E5	0.00
29)	L6 AR-1254-4	4.195					4.195	E5	0.00
30)	L6 AR-1254-5	4.767					4.767	E5	0.00
31)	L7 AR-1260-1	1.458	1.253	1.373	1.535	1.421	1.408	E5	7.45
32)	L7 AR-1260-2	1.770	1.528	1.656	1.881	1.852	1.738	E5	8.41
33)	L7 AR-1260-3	1.657	1.432	1.550	1.739	1.614	1.598	E5	7.22
34)	L7 AR-1260-4	1.067	0.936	0.993	1.156	1.087	1.048	E5	8.14
35)	L7 AR-1260-5	2.583	2.375	2.463	2.767	2.658	2.569	E5	6.03
36)	L8 AR-1262-1	2.111					2.111	E5	0.00
37)	L8 AR-1262-2	3.631					3.631	E5	0.00
38)	L8 AR-1262-3	1.402					1.402	E5	0.00
39)	L8 AR-1262-4	2.541					2.541	E5	0.00
40)	L8 AR-1262-5	1.040					1.040	E5	0.00
41)	L9 AR-1268-1	3.588					3.588	E5	0.00
42)	L9 AR-1268-2	3.244					3.244	E5	0.00
43)	L9 AR-1268-3	2.648					2.648	E5	0.00
44)	L9 AR-1268-4	1.063					1.063	E5	0.00
45)	L9 AR-1268-5	8.277					8.277	E5	0.00

(#) = Out of Range