

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030920\
 Data File : PQ046803.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2020 18:21
 Operator : AJ\MA
 Sample : L1841-40 5X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 WALL-FLASHING-B

Manual Integrations
 APPROVED

Sohil
 3/11/2020 11:00:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 05:44:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 10 04:56:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.179	4.324	77468312	35488276	4.145	3.934
2) SA Decachlor...	11.306	9.741	42783484	34014016	4.500	4.644
Target Compounds						
26) L6 AR-1254-1	7.407	6.487	9972218	5992907	18.817	15.715
27) L6 AR-1254-2	7.641	6.646	13158216	4129901	16.639m	12.627
28) L6 AR-1254-3	8.018	7.073	9256276	7769232	12.470m	14.344
29) L6 AR-1254-4	8.314	7.311	8185329	3555819	14.144m	10.821
30) L6 AR-1254-5	8.742	7.743	5543984	4759099	8.983m	10.293

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030920\
Data File : PQ046803.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2020 18:21
Operator : AJ\MA
Sample : L1841-40 5X
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
WALL-FLASHING-B


Manual Integrations
APPROVED

Sohil
3/11/2020 11:00:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 05:44:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030920.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 10 04:56:12 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 77468312
Conc: 4.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :
WALL-FLASHING-B


Manual Integrations
APPROVED

Sohil
3/11/2020 11:00:22 AM




#1 Tetrachloro-m-xylene

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 35488276
Conc: 3.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.306 min
Delta R.T.: 0.006 min
Response: 42783484
Conc: 4.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.741 min
Delta R.T.: 0.002 min
Response: 34014016
Conc: 4.64 ng/ml




#26 AR-1254-1

R.T.: 7.407 min
Delta R.T.: -0.001 min
Response: 9972218
Conc: 18.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
WALL-FLASHING-B


Manual Integrations
APPROVED

Sohil
3/11/2020 11:00:22 AM




#26 AR-1254-1

R.T.: 6.487 min
Delta R.T.: -0.001 min
Response: 5992907
Conc: 15.71 ng/ml




#27 AR-1254-2

R.T.: 7.641 min
Delta R.T.: 0.004 min
Response: 13158216
Conc: 16.64 ng/ml m



#27 AR-1254-2

R.T.: 6.646 min
Delta R.T.: 0.000 min
Response: 4129901
Conc: 12.63 ng/ml




#28 AR-1254-3

R.T.: 8.018 min
Delta R.T.: 0.000 min
Response: 9256276
Conc: 12.47 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
WALL-FLASHING-B


Manual Integrations
APPROVED

Sohil
3/11/2020 11:00:22 AM




#28 AR-1254-3

R.T.: 7.073 min
Delta R.T.: 0.000 min
Response: 7769232
Conc: 14.34 ng/ml




#29 AR-1254-4

R.T.: 8.314 min
Delta R.T.: 0.000 min
Response: 8185329
Conc: 14.14 ng/ml m



#29 AR-1254-4

R.T.: 7.311 min
Delta R.T.: 0.000 min
Response: 3555819
Conc: 10.82 ng/ml




#30 AR-1254-5

R.T.: 8.742 min
Delta R.T.: -0.002 min
Response: 5543984
Conc: 8.98 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
WALL-FLASHING-B

Manual Integrations
APPROVED

Sohil
3/11/2020 11:00:22 AM



#30 AR-1254-5

R.T.: 7.743 min
Delta R.T.: 0.000 min
Response: 4759099
Conc: 10.29 ng/ml