

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ031320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Mar 14 05:55:48 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ046956.D 1600 =PQ046958.D 800 =PQ046957.D
 200 =PQ046955.D 100 =PQ046954.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.905	1.685	1.630	1.738	1.635	1.719	E7	6.56
2) SA	Decachlorobiphenyl	0.934	0.866	0.917	1.054	1.104	0.975	E7	10.22
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	5.170	4.444	4.523	5.264	5.430	4.966	E5	9.08
37) L8	AR-1262-2	1.977	1.820	1.837	2.025	2.079	1.948	E6	5.89
38) L8	AR-1262-3	1.198	1.089	1.113	1.265	1.305	1.194	E6	7.83
39) L8	AR-1262-4	8.817	7.902	8.115	9.440	9.846	8.824	E5	9.44
40) L8	AR-1262-5	5.726	5.316	5.527	6.317	6.934	5.964	E5	11.04
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ046956.D 1600 =PQ046958.D 800 =PQ046957.D
 200 =PQ046955.D 100 =PQ046954.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.240	2.806	2.740	2.842	2.730	2.872	E6	7.35
2) SA	Decachlorobiphenyl	2.892	2.960	2.979	3.162	3.245	3.048	E6	4.88
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ031320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Mar 14 05:55:48 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ046956.D 1600 =PQ046958.D 800 =PQ046957.D
 200 =PQ046955.D 100 =PQ046954.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	1.345	1.223	1.205	1.316	1.324	1.283 E5	4.99
37) L8	AR-1262-2	5.108	4.817	4.796	4.992	4.944	4.931 E5	2.61
38) L8	AR-1262-3	1.875	1.767	1.765	1.914	1.959	1.856 E5	4.71
39) L8	AR-1262-4	3.603	3.492	3.464	3.633	3.641	3.566 E5	2.32
40) L8	AR-1262-5	1.614	1.610	1.611	1.726	1.730	1.658 E5	3.84
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range