

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ031320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Mar 14 06:48:35 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ046961.D 1600 =PQ046963.D 800 =PQ046962.D
 200 =PQ046960.D 100 =PQ046959.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.738	1.613	1.556	1.822	1.739	1.694	E7	6.32
2) SA	Decachlorobiphenyl	1.711	1.536	1.585	1.844	1.919	1.719	E7	9.50
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	2.091	1.947	1.956	2.221	2.304	2.104	E6	7.53
42) L9	AR-1268-2	1.774	1.645	1.644	1.903	1.963	1.786	E6	8.16
43) L9	AR-1268-3	1.460	1.339	1.353	1.568	1.644	1.473	E6	9.02
44) L9	AR-1268-4	6.422	5.760	5.907	6.900	7.201	6.438	E5	9.63
45) L9	AR-1268-5	4.086	3.787	3.855	4.323	4.432	4.097	E6	6.89

Signal #2 Calibration Files

400 =PQ046961.D 1600 =PQ046963.D 800 =PQ046962.D
 200 =PQ046960.D 100 =PQ046959.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.942	2.709	2.646	3.058	3.095	2.890	E6	7.03
2) SA	Decachlorobiphenyl	5.573	5.466	5.401	5.740	5.875	5.611	E6	3.48
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ031320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Mar 14 06:48:35 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ046961.D 1600 =PQ046963.D 800 =PQ046962.D
 200 =PQ046960.D 100 =PQ046959.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	5.435	5.295	5.179	5.650	5.707	5.453 E5	4.13
42) L9	AR-1268-2	4.960	4.899	4.764	5.128	5.152	4.981 E5	3.26
43) L9	AR-1268-3	4.045	4.023	3.888	4.189	4.231	4.075 E5	3.38
44) L9	AR-1268-4	1.803	1.754	1.715	1.883	1.989	1.829 E5	5.98
45) L9	AR-1268-5	1.359	1.312	1.314	1.381	1.371	1.348 E6	2.40

(#) = Out of Range