



Suggestion Box MSDS RICCA MSDS J.T.Baker
MSDS emdmillipore MSDS restek MSDS CHEMTECH

Welcome Sakhvala
| Logout

Search :

Dashboard | Edd | Efolder | Fax | Genchem | Projecttrack | Qaqc | Quanttrack | Report | Sequence | Utility |

[Home](#) >> [Sequence Module](#) >> [File Info](#)

Sequence Module :: File Info

Navigation Bar

Method : 608.3 PCB

[FileInfo Form](#)

[Get Test Info](#)

[Analysis RunLog](#)

[Sequence Info](#)

[Maintenance](#)

[Audit Summary Report](#)

[Sequence Process](#)

Data For Sequence pq032620 **CHEMTECH-NJ Lab Sequence**
Rule For Ms/Msd of EPH

[Sequence Transfer](#)

[Review Raw Data](#)

[Non Conformance Summary](#)

No Review & No Supervised

Reviewed & No Supervised

Reviewed & Supervised

SampleID	Data File	Date & Time	ClientId	Test	QCType	Dilution.	Comments	Status	System Comments
I.BLK	PQ047068.D	26 Mar 2020 11:19	I.BLK	PCB	IBLK	1	--	OK	UPDATE STATUS
AR1660CCC500	PQ047069.D	26 Mar 2020 11:35	AR1660CCC500	PCB	CCAL500	1	--	OK	UPDATE STATUS
PB127794BL	PQ047070.D	26 Mar 2020 11:54	PB127794BL	PCB	MB	1	--	OK	UPDATE STATUS
PB127794BS	PQ047071.D	26 Mar 2020 12:11	PB127794BS	PCB	LCS	1	--	OK	UPDATE STATUS
L2017-01	PQ047072.D	26 Mar 2020 12:27	FE-SB-32-(8.5-9)	PCB	SAM	1	--	OK	UPDATE STATUS
L2017-02	PQ047073.D	26 Mar 2020 12:43	FE-SB-35-(7-7.5)	PCB	SAM	1	--	OK	UPDATE STATUS
L2017-03	PQ047074.D	26 Mar 2020 12:59	FE-SB-29-(5.5-6)	PCB	SAM	1	--	OK	UPDATE STATUS
L2017-03MS	PQ047075.D	26 Mar 2020 13:16	FE-SB-29-(5.5-6)MS	PCB	MS	1	--	OK	UPDATE STATUS
L2017-03MSD	PQ047076.D	26 Mar 2020 13:32	FE-SB-29-(5.5-6)MSD	PCB	MSD	1	--	OK	UPDATE STATUS
L2028-01	PQ047077.D	26 Mar 2020 13:48	SB-30-(8-8.5)	PCB	SAM	1	--	OK	UPDATE STATUS
		26					AR1254 &		

L2028-02	PO047078.D	Mar 2020 14:04	SB-31-(5.5-6)	PCB	SAM	1	AR1260 hits, Plz confirm hit	OK	UPDATE STATUS
I.BLK	PO047079.D	26 Mar 2020 14:21	I.BLK	PCB	IBLK	1	--	OK	UPDATE STATUS
AR1660CCC500	PO047080.D	26 Mar 2020 15:59	AR1660CCC500	PCB	CCAL500	1	--	OK	UPDATE STATUS
L1753-09	PO047081.D	26 Mar 2020 16:40	CL-02-032520	PCB	SAM	1	End CCC fail	NotOK	UPDATE STATUS
L2031-18	PO047082.D	26 Mar 2020 16:56	DEBRIS-COMP	PCB	SAM	1	DCB failing low in both coloumn,End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2031-20	PO047083.D	26 Mar 2020 17:13	SOIL-COMP	PCB	SAM	1	AR1242 Hits, Need 40x dilution, End CCC fail	NotOK	Surr, OR Fail UPDATE STATUS
L2035-01	PO047084.D	26 Mar 2020 17:29	FK-GF-02-054-A	PCB	SAM	1	DCB failing low in both coloumn, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2035-03	PO047085.D	26 Mar 2020 17:45	FK-GF-02-054-B	PCB	SAM	1	DCB low in 1st column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-01	PO047086.D	26 Mar 2020 18:02	03-24-A	PCB	SAM	2	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-02	PO047087.D	26 Mar 2020 18:18	03-24-B	PCB	SAM	2	End CCC fail	NotOK	UPDATE STATUS
L2027-03	PO047088.D	26 Mar 2020 18:34	03-24-C	PCB	SAM	2	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-04	PO047089.D	26 Mar 2020 18:51	03-24-D	PCB	SAM	2	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-05	PO047090.D	26 Mar 2020 19:07	03-24-E	PCB	SAM	2	DCB low in 2nd column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-06	PO047091.D	26 Mar 2020 19:23	03-24-F	PCB	SAM	2	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-07	PO047092.D	26 Mar 2020 19:39	03-24-G	PCB	SAM	2	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2027-08	PO047093.D	26 Mar 2020 19:56	03-24-H	PCB	SAM	2	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS
L2042-03	PO047094.D	26 Mar 2020 20:12	SB-03-5.5-6.0	PCB	SAM	1	AR1254 hit, Confirm hit, End CCC fail	NotOK	UPDATE STATUS
L2042-04	PO047095.D	26 Mar 2020 20:28	SB-04-5.5-6.0	PCB	SAM	1	DCB low in both column, End CCC fail	NotOK	Surr Fail UPDATE STATUS

L2034-01	PQ047096.D	26 Mar 2020 20:45	FK-BPM-04-008-ABCD	PCB	SAM	1	End CCC fail	NotOK	UPDATE STATUS
L2034-01MS	PQ047097.D	26 Mar 2020 21:01	FK-BPM-04-008-ABCDMS	PCB	MS	1	End CCC fail	NotOK	UPDATE STATUS
L2034-01MSD	PQ047098.D	26 Mar 2020 21:17	FK-BPM-04-008-ABCDMSD	PCB	MSD	1	End CCC fail	NotOK	UPDATE STATUS
PB127820BL	PQ047099.D	26 Mar 2020 21:34	PB127820BL	PCB	MB	1	End CCC fail	NotOK	UPDATE STATUS
PB127820BS	PQ047100.D	26 Mar 2020 21:50	PB127820BS	PCB	LCS	1	End CCC fail	NotOK	UPDATE STATUS
I.BLK	PQ047101.D	26 Mar 2020 22:06	I.BLK	PCB	IBLK	1	IBLK is Looks like hexane	NotOK	Surr Fail UPDATE STATUS
AR1660CCC500	PQ047102.D	26 Mar 2020 22:22	AR1660CCC500	PCB	CCAL500	1	CCC is Looks like hexane	NotOK	CCAL Fail UPDATE STATUS
Read Files		Update		All Manual Integration Report					

[Read QUANT File](#)[EPH Sequence Transfer](#)[Update ClientInfo](#)[Trasfer PH](#)