

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ033020\
 Data File : PQ047129.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2020 13:34
 Operator : AJ\MA
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 14:15:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ033020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 30 14:15:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.108	4.264	964.0E6	128.6E6	50.000	50.000
2) SA Decachlor...	11.182	9.651	422.2E6	111.6E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	7.332	6.416	334.5E6	71727170	500.000	500.000
27) L6 AR-1254-2	7.562	6.575	506.8E6	61885046	500.000	500.000
28) L6 AR-1254-3	7.943	7.000	491.6E6	101.2E6	500.000	500.000
29) L6 AR-1254-4	8.237	7.239	381.5E6	64247035	500.000	500.000
30) L6 AR-1254-5	8.668	7.671	337.7E6	86002796	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ033020\
Data File : PQ047129.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2020 13:34
Operator : AJ\MA
Sample : AR1254ICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 14:15:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ033020.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 30 14:15:26 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

