

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ033020\
 Data File : PQ047130.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2020 13:51
 Operator : AJ\MA
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 14:25:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ033020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 30 14:25:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.107	4.264	891.6E6	118.4E6	50.000	50.000
2) SA Decachlor...	11.182	9.651	431.1E6	121.9E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.743	7.670	447.9E6	43478019	500.000	500.000
37) L8 AR-1262-2	9.310	8.221	678.2E6	164.1E6	500.000	500.000
38) L8 AR-1262-3	9.632	8.510	432.1E6	66172540	500.000	500.000
39) L8 AR-1262-4	9.724	8.575	321.3E6	118.0E6	500.000	500.000
40) L8 AR-1262-5	10.406	9.082	219.6E6	55341927	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ033020\
Data File : PQ047130.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2020 13:51
Operator : AJ\MA
Sample : AR1262ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 14:25:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ033020.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 30 14:25:20 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

