

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041620\
 Data File : PQ047427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2020 16:54
 Operator : AJ\MA
 Sample : PQ041620ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 08:23:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 17 08:20:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	50.000	45.908	8.2	90	0.00
2 SA Decachlorobiphenyl	50.000	51.456	-2.9	106	0.00
3 L1 AR-1016-1	500.000	463.479	7.3	95	0.00
4 L1 AR-1016-2	500.000	469.966	6.0	96	0.00
5 L1 AR-1016-3	500.000	470.350	5.9	97	0.00
6 L1 AR-1016-4	500.000	475.850	4.8	98	0.00
7 L1 AR-1016-5	500.000	480.930	3.8	98	0.00
31 L7 AR-1260-1	500.000	499.539	0.1	101	0.00
32 L7 AR-1260-2	500.000	497.703	0.5	101	0.00
33 L7 AR-1260-3	500.000	512.124	-2.4	105	0.00
34 L7 AR-1260-4	500.000	508.935	-1.8	105	0.00
35 L7 AR-1260-5	500.000	512.321	-2.5	103	0.00

Signal #2

1 SA Tetrachloro-m-xylene	50.000	44.589	10.8	86	0.00
2 SA Decachlorobiphenyl	50.000	50.474	-0.9	104	0.00
3 L1 AR-1016-1	500.000	445.049	11.0	90	0.00
4 L1 AR-1016-2	500.000	448.182	10.4	90	0.00
5 L1 AR-1016-3	500.000	455.928	8.8	91	0.00
6 L1 AR-1016-4	500.000	456.188	8.8	91	0.00
7 L1 AR-1016-5	500.000	446.269	10.7	89	0.00
31 L7 AR-1260-1	500.000	461.096	7.8	92	0.00
32 L7 AR-1260-2	500.000	464.772	7.0	94	0.00
33 L7 AR-1260-3	500.000	467.483	6.5	95	0.00
34 L7 AR-1260-4	500.000	479.247	4.2	96	0.00
35 L7 AR-1260-5	500.000	496.338	0.7	99	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	500.000	0.000	100.0#	0	-5.35#
9 L2 AR-1221-2	500.000	0.000	100.0#	0	-5.45#
10 L2 AR-1221-3	500.000	0.000	100.0#	0	-5.54#
11 L3 AR-1232-1	500.000	0.000	100.0#	0	-5.54#
12 L3 AR-1232-2	500.000	0.000	100.0#	0	-6.13#
13 L3 AR-1232-3	500.000	0.000	100.0#	0	-6.45#
14 L3 AR-1232-4	500.000	0.000	100.0#	0	-6.62#
15 L3 AR-1232-5	500.000	0.000	100.0#	0	-6.72#
16 L4 AR-1242-1	500.000	0.000	100.0#	0	-6.42#
17 L4 AR-1242-2	500.000	0.000	100.0#	0	-6.45#
18 L4 AR-1242-3	500.000	0.000	100.0#	0	-6.52#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041620\
 Data File : PQ047427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2020 16:54
 Operator : AJ\MA
 Sample : PQ041620ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 08:23:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 17 08:20:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	500.000	0.000	100.0#	0	-6.62#
20 L4 AR-1242-5	500.000	0.000	100.0#	0	-7.40#
21 L5 AR-1248-1	500.000	0.000	100.0#	0	-6.42#
22 L5 AR-1248-2	500.000	0.000	100.0#	0	-6.72#
23 L5 AR-1248-3	500.000	0.000	100.0#	0	-6.94#
24 L5 AR-1248-4	500.000	0.000	100.0#	0	-7.36#
25 L5 AR-1248-5	500.000	0.000	100.0#	0	-7.40#
26 L6 AR-1254-1	500.000	0.000	100.0#	0	-7.34#
27 L6 AR-1254-2	500.000	0.000	100.0#	0	-7.56#
28 L6 AR-1254-3	500.000	0.000	100.0#	0	-7.95#
29 L6 AR-1254-4	500.000	0.000	100.0#	0	-8.24#
30 L6 AR-1254-5	500.000	0.000	100.0#	0	-8.67#
36 L8 AR-1262-1	500.000	0.000	100.0#	0	-8.76#
37 L8 AR-1262-2	500.000	0.000	100.0#	0	-9.33#
38 L8 AR-1262-3	500.000	0.000	100.0#	0	-9.65#
39 L8 AR-1262-4	500.000	0.000	100.0#	0	-9.75#
40 L8 AR-1262-5	500.000	0.000	100.0#	0	-10.43#
41 L9 AR-1268-1	500.000	0.000	100.0#	0	-9.64#
42 L9 AR-1268-2	500.000	0.000	100.0#	0	-9.74#
43 L9 AR-1268-3	500.000	0.000	100.0#	0	-9.97#
44 L9 AR-1268-4	500.000	0.000	100.0#	0	-10.42#
45 L9 AR-1268-5	500.000	0.000	100.0#	0	-10.84#

Signal #2

8 L2 AR-1221-1	500.000	0.000	100.0#	0	-4.52#
9 L2 AR-1221-2	500.000	0.000	100.0#	0	-4.63#
10 L2 AR-1221-3	500.000	0.000	100.0#	0	-4.72#
11 L3 AR-1232-1	500.000	0.000	100.0#	0	-4.72#
12 L3 AR-1232-2	500.000	0.000	100.0#	0	-5.56#
13 L3 AR-1232-3	500.000	0.000	100.0#	0	-5.75#
14 L3 AR-1232-4	500.000	0.000	100.0#	0	-5.85#
15 L3 AR-1232-5	500.000	0.000	100.0#	0	-6.03#
16 L4 AR-1242-1	500.000	0.000	100.0#	0	-5.54#
17 L4 AR-1242-2	500.000	0.000	100.0#	0	-5.56#
18 L4 AR-1242-3	500.000	0.000	100.0#	0	-5.75#
19 L4 AR-1242-4	500.000	0.000	100.0#	0	-5.85#
20 L4 AR-1242-5	500.000	0.000	100.0#	0	-6.41#
21 L5 AR-1248-1	500.000	0.000	100.0#	0	-5.54#
22 L5 AR-1248-2	500.000	0.000	100.0#	0	-5.80#
23 L5 AR-1248-3	500.000	0.000	100.0#	0	-5.84#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ041620\
 Data File : PQ047427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2020 16:54
 Operator : AJ\MA
 Sample : PQ041620ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 08:23:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ041620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 17 08:20:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	500.000	0.000	100.0#	0	-6.03#
25 L5 AR-1248-5	500.000	0.000	100.0#	0	-6.46#
26 L6 AR-1254-1	500.000	0.000	100.0#	0	-6.41#
27 L6 AR-1254-2	500.000	0.000	100.0#	0	-6.57#
28 L6 AR-1254-3	500.000	0.000	100.0#	0	-7.00#
29 L6 AR-1254-4	500.000	0.000	100.0#	0	-7.24#
30 L6 AR-1254-5	500.000	0.000	100.0#	0	-7.67#
36 L8 AR-1262-1	500.000	0.000	100.0#	0	-7.67#
37 L8 AR-1262-2	500.000	0.000	100.0#	0	-8.22#
38 L8 AR-1262-3	500.000	0.000	100.0#	0	-8.51#
39 L8 AR-1262-4	500.000	0.000	100.0#	0	-8.58#
40 L8 AR-1262-5	500.000	0.000	100.0#	0	-9.08#
41 L9 AR-1268-1	500.000	0.000	100.0#	0	-8.51#
42 L9 AR-1268-2	500.000	0.000	100.0#	0	-8.57#
43 L9 AR-1268-3	500.000	0.000	100.0#	0	-8.78#
44 L9 AR-1268-4	500.000	0.000	100.0#	0	-9.08#
45 L9 AR-1268-5	500.000	0.000	100.0#	0	-9.38#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0