

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051119\
 Data File : PQ039663.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2019 20:56
 Operator : AJ\MA
 Sample : K2765-15
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P026-SS005-0612-01

Manual Integrations
 APPROVED

Ankita
 5/15/2019 2:01:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 00:27:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 10 01:55:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.655	4.761	135.8E6	84574046	23.758m	23.046
2) SA Decachlor...	12.027	10.510	105.5E6	49453445	16.322m	14.925
Target Compounds						
31) L7 AR-1260-1	8.872	7.920	60785174	43835517	238.389	247.520
32) L7 AR-1260-2	9.142	8.125	118.8E6	86839108	381.035	395.764
33) L7 AR-1260-3	9.517	8.289	86522001	43974525	346.139	212.636 #
34) L7 AR-1260-4	9.757	8.787	88078101	31680277	305.281	181.561 #
35) L7 AR-1260-5	10.104	9.040	138.4E6	107.2E6	234.187	252.900

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051119\
Data File : PQ039663.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2019 20:56
Operator : AJ\MA
Sample : K2765-15
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
P026-SS005-0612-01

Manual Integrations
APPROVED

Ankita
5/15/2019 2:01:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 00:27:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ050919.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 10 01:55:02 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

