

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051820CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 19 01:42:45 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ047788.D 1600 =PQ047790.D 800 =PQ047789.D
 200 =PQ047787.D 100 =PQ047786.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	7.581	7.433	7.641	7.610	8.490	7.751	E6	5.43
2) SA	Decachlorobiphenyl	7.333	6.675	7.008	7.842	8.339	7.439	E6	8.90
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	1.985	1.808	1.907	2.160	2.266	2.025	E5	9.21
37) L8	AR-1262-2	8.962	8.783	8.921	9.375	9.925	9.193	E5	5.06
38) L8	AR-1262-3	4.151	3.795	3.969	4.457	4.920	4.259	E5	10.42
39) L8	AR-1262-4	4.810	4.454	4.642	5.129	5.436	4.894	E5	8.00
40) L8	AR-1262-5	3.569	3.282	3.442	3.788	4.020	3.620	E5	8.01
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ047788.D 1600 =PQ047790.D 800 =PQ047789.D
 200 =PQ047787.D 100 =PQ047786.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.142	3.262	3.260	3.127	3.451	3.248	E6	3.99
2) SA	Decachlorobiphenyl	3.487	3.334	3.409	3.699	3.979	3.582	E6	7.27
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051820CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 19 01:42:45 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ047788.D 1600 =PQ047790.D 800 =PQ047789.D
 200 =PQ047787.D 100 =PQ047786.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	1.383	1.336	1.365	1.456	1.555	1.419 E5	6.21
37) L8	AR-1262-2	5.045	5.161	5.177	5.122	5.278	5.156 E5	1.64
38) L8	AR-1262-3	1.975	1.917	2.058	2.076	2.216	2.048 E5	5.55
39) L8	AR-1262-4	3.699	3.738	3.760	3.823	4.010	3.806 E5	3.23
40) L8	AR-1262-5	1.782	1.731	1.763	1.875	2.001	1.831 E5	5.97
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range