

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051820CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 19 02:22:00 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ047793.D 1600 =PQ047795.D 800 =PQ047794.D
 200 =PQ047792.D 100 =PQ047791.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	7.765	7.570	7.497	7.977	8.139	7.790	E6	3.47
2) SA	Decachlorobiphenyl	1.262	1.199	1.226	1.375	1.446	1.302	E7	8.08
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	1.018	1.005	1.010	1.099	1.151	1.057	E6	6.16
42) L9	AR-1268-2	0.967	0.952	0.958	1.045	1.091	1.003	E6	6.18
43) L9	AR-1268-3	8.153	8.029	8.075	8.809	9.247	8.463	E5	6.39
44) L9	AR-1268-4	3.845	3.587	3.732	4.158	4.364	3.937	E5	8.07
45) L9	AR-1268-5	2.809	2.832	2.814	2.960	3.064	2.896	E6	3.88

Signal #2 Calibration Files

400 =PQ047793.D 1600 =PQ047795.D 800 =PQ047794.D
 200 =PQ047792.D 100 =PQ047791.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.255	3.324	3.244	3.263	3.284	3.274	E6	0.96
2) SA	Decachlorobiphenyl	6.153	6.131	6.136	6.557	6.878	6.371	E6	5.28
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ051820CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 19 02:22:00 2020
 Response Via : Initial Calibration

Calibration Files

400 =PQ047793.D 1600 =PQ047795.D 800 =PQ047794.D
 200 =PQ047792.D 100 =PQ047791.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	5.618	5.738	5.732	5.923	6.080	5.818 E5	3.14
42) L9	AR-1268-2	5.157	5.326	5.272	5.410	5.560	5.345 E5	2.83
43) L9	AR-1268-3	4.253	4.386	4.336	4.480	4.596	4.410 E5	3.00
44) L9	AR-1268-4	1.912	1.903	1.906	2.056	2.113	1.978 E5	5.01
45) L9	AR-1268-5	1.446	1.460	1.472	1.488	1.492	1.472 E6	1.31

(#) = Out of Range