

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051820\
 Data File : PQ047785.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 May 2020 18:22
 Operator : AJ\MA
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 19:07:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051820CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 18 18:41:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.328	4.299	592.4E6	258.1E6	76.973	81.018
2) SA Decachlor...	11.632	9.684	1068.0E6	528.6E6	144.613	151.704
Target Compounds						
26) L6 AR-1254-1	7.574	6.445	401.2E6	289.2E6	1418.396	1507.428
27) L6 AR-1254-2	7.803	6.603	629.6E6	250.9E6	1434.335	1487.127
28) L6 AR-1254-3	8.189	7.026	698.0E6	433.8E6	1477.697	1547.857
29) L6 AR-1254-4	8.485	7.264	531.5E6	279.4E6	1464.114	1532.075
30) L6 AR-1254-5	8.918	7.696	583.2E6	389.0E6	1476.039	1544.485

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ051820\
Data File : PQ047785.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 May 2020 18:22
Operator : AJ\MA
Sample : AR1254ICC1600
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1254501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 19:07:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ051820CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 18 18:41:52 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

