

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052720\
 Data File : PQ047846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2020 17:26
 Operator : AJ\MA
 Sample : PQ052720ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 17:40:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 27 17:33:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	50.000	52.802	-5.6	107	0.00
2 SA Decachlorobiphenyl	50.000	50.274	-0.5	100	0.00
3 L1 AR-1016-1	500.000	505.355	-1.1	104	0.00
4 L1 AR-1016-2	500.000	509.449	-1.9	104	0.00
5 L1 AR-1016-3	500.000	499.877	0.0	105	0.00
6 L1 AR-1016-4	500.000	501.495	-0.3	104	0.00
7 L1 AR-1016-5	500.000	497.721	0.5	105	0.00
31 L7 AR-1260-1	500.000	482.995	3.4	101	0.00
32 L7 AR-1260-2	500.000	504.464	-0.9	103	0.00
33 L7 AR-1260-3	500.000	501.215	-0.2	101	0.00
34 L7 AR-1260-4	500.000	506.745	-1.3	100	0.00
35 L7 AR-1260-5	500.000	508.817	-1.8	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	50.000	52.484	-5.0	105	0.00
2 SA Decachlorobiphenyl	50.000	50.579	-1.2	101	0.00
3 L1 AR-1016-1	500.000	511.951	-2.4	105	0.00
4 L1 AR-1016-2	500.000	503.684	-0.7	101	0.00
5 L1 AR-1016-3	500.000	513.926	-2.8	103	0.00
6 L1 AR-1016-4	500.000	516.309	-3.3	104	0.00
7 L1 AR-1016-5	500.000	509.389	-1.9	104	0.00
31 L7 AR-1260-1	500.000	495.737	0.9	102	0.00
32 L7 AR-1260-2	500.000	508.679	-1.7	102	0.00
33 L7 AR-1260-3	500.000	512.880	-2.6	102	0.00
34 L7 AR-1260-4	500.000	503.369	-0.7	102	0.00
35 L7 AR-1260-5	500.000	520.100	-4.0	101	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	500.000	0.000	100.0#	0	-5.57#
9 L2 AR-1221-2	500.000	0.000	100.0#	0	-5.67#
10 L2 AR-1221-3	500.000	0.000	100.0#	0	-5.76#
11 L3 AR-1232-1	500.000	0.000	100.0#	0	-5.76#
12 L3 AR-1232-2	500.000	0.000	100.0#	0	-6.36#
13 L3 AR-1232-3	500.000	0.000	100.0#	0	-6.68#
14 L3 AR-1232-4	500.000	0.000	100.0#	0	-6.85#
15 L3 AR-1232-5	500.000	0.000	100.0#	0	-6.95#
16 L4 AR-1242-1	500.000	0.000	100.0#	0	-6.65#
17 L4 AR-1242-2	500.000	0.000	100.0#	0	-6.68#
18 L4 AR-1242-3	500.000	0.000	100.0#	0	-6.75#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052720\
 Data File : PQ047846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2020 17:26
 Operator : AJ\MA
 Sample : PQ052720ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 17:40:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 27 17:33:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	500.000	0.000	100.0#	0	-6.85#
20 L4 AR-1242-5	500.000	0.000	100.0#	0	-7.64#
21 L5 AR-1248-1	500.000	0.000	100.0#	0	-6.65#
22 L5 AR-1248-2	500.000	0.000	100.0#	0	-6.95#
23 L5 AR-1248-3	500.000	0.000	100.0#	0	-7.17#
24 L5 AR-1248-4	500.000	0.000	100.0#	0	-7.60#
25 L5 AR-1248-5	500.000	0.000	100.0#	0	-7.64#
26 L6 AR-1254-1	500.000	0.000	100.0#	0	-7.57#
27 L6 AR-1254-2	500.000	0.000	100.0#	0	-7.80#
28 L6 AR-1254-3	500.000	0.000	100.0#	0	-8.18#
29 L6 AR-1254-4	500.000	0.000	100.0#	0	-8.48#
30 L6 AR-1254-5	500.000	0.000	100.0#	0	-8.91#
36 L8 AR-1262-1	500.000	0.000	100.0#	0	-8.99#
37 L8 AR-1262-2	500.000	0.000	100.0#	0	-9.59#
38 L8 AR-1262-3	500.000	0.000	100.0#	0	-9.93#
39 L8 AR-1262-4	500.000	0.000	100.0#	0	-10.03#
40 L8 AR-1262-5	500.000	0.000	100.0#	0	-10.78#
41 L9 AR-1268-1	500.000	0.000	100.0#	0	-9.94#
42 L9 AR-1268-2	500.000	0.000	100.0#	0	-10.04#
43 L9 AR-1268-3	500.000	0.000	100.0#	0	-10.30#
44 L9 AR-1268-4	500.000	0.000	100.0#	0	-10.78#
45 L9 AR-1268-5	500.000	0.000	100.0#	0	-11.25#

Signal #2

8 L2 AR-1221-1	500.000	0.000	100.0#	0	-4.56#
9 L2 AR-1221-2	500.000	0.000	100.0#	0	-4.66#
10 L2 AR-1221-3	500.000	0.000	100.0#	0	-4.75#
11 L3 AR-1232-1	500.000	0.000	100.0#	0	-4.75#
12 L3 AR-1232-2	500.000	0.000	100.0#	0	-5.59#
13 L3 AR-1232-3	500.000	0.000	100.0#	0	-5.78#
14 L3 AR-1232-4	500.000	0.000	100.0#	0	-5.88#
15 L3 AR-1232-5	500.000	0.000	100.0#	0	-6.06#
16 L4 AR-1242-1	500.000	0.000	100.0#	0	-5.57#
17 L4 AR-1242-2	500.000	0.000	100.0#	0	-5.59#
18 L4 AR-1242-3	500.000	0.000	100.0#	0	-5.78#
19 L4 AR-1242-4	500.000	0.000	100.0#	0	-5.88#
20 L4 AR-1242-5	500.000	0.000	100.0#	0	-6.44#
21 L5 AR-1248-1	500.000	0.000	100.0#	0	-5.57#
22 L5 AR-1248-2	500.000	0.000	100.0#	0	-5.83#
23 L5 AR-1248-3	500.000	0.000	100.0#	0	-5.88#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052720\
 Data File : PQ047846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2020 17:26
 Operator : AJ\MA
 Sample : PQ052720ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 17:40:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 27 17:33:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	500.000	0.000	100.0#	0	-6.06#
25 L5 AR-1248-5	500.000	0.000	100.0#	0	-6.48#
26 L6 AR-1254-1	500.000	0.000	100.0#	0	-6.44#
27 L6 AR-1254-2	500.000	0.000	100.0#	0	-6.60#
28 L6 AR-1254-3	500.000	0.000	100.0#	0	-7.02#
29 L6 AR-1254-4	500.000	0.000	100.0#	0	-7.26#
30 L6 AR-1254-5	500.000	0.000	100.0#	0	-7.69#
36 L8 AR-1262-1	500.000	0.000	100.0#	0	-7.69#
37 L8 AR-1262-2	500.000	0.000	100.0#	0	-8.24#
38 L8 AR-1262-3	500.000	0.000	100.0#	0	-8.53#
39 L8 AR-1262-4	500.000	0.000	100.0#	0	-8.60#
40 L8 AR-1262-5	500.000	0.000	100.0#	0	-9.10#
41 L9 AR-1268-1	500.000	0.000	100.0#	0	-8.53#
42 L9 AR-1268-2	500.000	0.000	100.0#	0	-8.60#
43 L9 AR-1268-3	500.000	0.000	100.0#	0	-8.81#
44 L9 AR-1268-4	500.000	0.000	100.0#	0	-9.10#
45 L9 AR-1268-5	500.000	0.000	100.0#	0	-9.41#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0