

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061320\
 Data File : PQ048396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2020 13:47
 Operator : AJ\MA
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 16:20:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 12 16:17:22 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.295	4.289	322.6E6	151.5E6	50.000	50.000
2)	SA Decachlor...	11.566	9.660	532.1E6	259.6E6	50.000	50.000
Target Compounds							
16)	L4 AR-1242-1	6.625	5.555	102.4E6	50885462	500.000	500.000
17)	L4 AR-1242-2	6.649	5.576	139.6E6	69371251	500.000	500.000
18)	L4 AR-1242-3	6.716	5.769	87066387	37148341	500.000	500.000
19)	L4 AR-1242-4	6.824	5.864	72314904	37538020	500.000	500.000
20)	L4 AR-1242-5	7.610	6.429	84843404	52273308	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061320\
Data File : PQ048396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2020 13:47
Operator : AJ\MA
Sample : AR1242ICC500
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 16:20:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061320.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 12 16:17:22 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

