

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062518\
 Data File : PQ030079.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2018 12:50
 Operator : UA/SM
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 03:10:15 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\PQ062518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 26 03:10:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.439	3.702	2655.1E6	582.1E6	48.136	49.288
2) SA Decachlor...	10.123	8.677	1748.8E6	668.6E6	44.518	46.071
Target Compounds						
36) L8 AR-1262-1	7.189	6.248	1039.7E6	296.8E6	500.000	500.000
37) L8 AR-1262-2	7.443	6.433	1155.5E6	357.7E6	500.000	500.000
38) L8 AR-1262-3	7.798	6.797	1539.1E6	533.9E6	500.000	500.000
39) L8 AR-1262-4	8.024	7.056	1420.9E6	467.9E6	500.000	500.000
40) L8 AR-1262-5	8.341	7.293	2690.3E6	1013.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062518\
Data File : PQ030079.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2018 12:50
Operator : UA/SM
Sample : AR1262ICC500
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 03:10:15 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\PQ062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 26 03:10:10 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

