

Method Path : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\
 Method File : PQ062518.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jun 26 03:14:15 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030071.D 1000 =PQ030069.D 750 =PQ030070.D
 250 =PQ030072.D 50 =PQ030073.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.937	5.660	5.440	5.756	5.413	5.641	E7	3.90
2) SA	Decachlorobiphenyl	4.195	3.837	4.274	4.104	3.929	4.068	E7	4.48
3) L1	AR-1016-1	1.049	0.973	0.970	0.968	0.866	0.965	E6	6.75
4) L1	AR-1016-2	1.330	1.350	1.199	1.184	0.907	1.194	E6	14.83
5) L1	AR-1016-3	1.986	1.880	1.676	1.707	1.582	1.766	E6	9.26
6) L1	AR-1016-4	1.706	1.360	1.465	1.421	1.167	1.424	E6	13.67
7) L1	AR-1016-5	1.615	1.575	1.417	1.363	1.235	1.441	E6	10.83
8) L2	AR-1221-1	8.163					8.163	E5	0.00
9) L2	AR-1221-2	6.123					6.123	E5	0.00
10) L2	AR-1221-3	1.725					1.725	E6	0.00
11) L3	AR-1232-1	1.147					1.147	E6	0.00
12) L3	AR-1232-2	5.586					5.586	E5	0.00
13) L3	AR-1232-3	9.245					9.245	E5	0.00
14) L3	AR-1232-4	7.758					7.758	E5	0.00
15) L3	AR-1232-5	6.643					6.643	E5	0.00
16) L4	AR-1242-1	1.069					1.069	E6	0.00
17) L4	AR-1242-2	1.625					1.625	E6	0.00
18) L4	AR-1242-3	1.536					1.536	E6	0.00
19) L4	AR-1242-4	1.261					1.261	E6	0.00
20) L4	AR-1242-5	1.367					1.367	E6	0.00
21) L5	AR-1248-1	1.787					1.787	E6	0.00
22) L5	AR-1248-2	1.572					1.572	E6	0.00
23) L5	AR-1248-3	1.983					1.983	E6	0.00
24) L5	AR-1248-4	2.021					2.021	E6	0.00
25) L5	AR-1248-5	2.007					2.007	E6	0.00
26) L6	AR-1254-1	2.207					2.207	E6	0.00
27) L6	AR-1254-2	3.325					3.325	E6	0.00
28) L6	AR-1254-3	3.568					3.568	E6	0.00
29) L6	AR-1254-4	2.669					2.669	E6	0.00
30) L6	AR-1254-5	3.154					3.154	E6	0.00
31) L7	AR-1260-1	2.736	2.512	2.466	2.686	3.048	2.690	E6	8.56
32) L7	AR-1260-2	3.076	2.637	2.857	3.021	3.109	2.940	E6	6.64
33) L7	AR-1260-3	4.138	3.459	3.858	3.567	4.114	3.827	E6	8.08
34) L7	AR-1260-4	2.529	2.189	2.377	2.525	2.412	2.406	E6	5.78
35) L7	AR-1260-5	4.968	4.632	4.914	4.814	4.302	4.726	E6	5.70
36) L8	AR-1262-1	2.079					2.079	E6	0.00
37) L8	AR-1262-2	2.311					2.311	E6	0.00
38) L8	AR-1262-3	3.078					3.078	E6	0.00
39) L8	AR-1262-4	2.842					2.842	E6	0.00
40) L8	AR-1262-5	5.381					5.381	E6	0.00
41) L9	AR-1268-1	5.185					5.185	E6	0.00
42) L9	AR-1268-2	4.756					4.756	E6	0.00
43) L9	AR-1268-3	3.941					3.941	E6	0.00
44) L9	AR-1268-4	1.853					1.853	E6	0.00
45) L9	AR-1268-5	1.273					1.273	E7	0.00

Signal #2 Calibration Files

500 =PQ030071.D 1000 =PQ030069.D 750 =PQ030070.D
 250 =PQ030072.D 50 =PQ030073.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.298	1.053	1.314	1.261	1.113	1.208	E7	9.74
2) SA	Decachlorobiphenyl	1.564	1.284	1.571	1.575	1.489	1.497	E7	8.27
3) L1	AR-1016-1	3.235	2.623	3.323	3.300	2.980	3.092	E5	9.57
4) L1	AR-1016-2	3.302	2.701	3.330	3.268	3.057	3.132	E5	8.42
5) L1	AR-1016-3	2.514	2.053	2.520	2.525	2.195	2.361	E5	9.41

Method Path : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\
 Method File : PQ062518.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jun 26 03:14:15 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030071.D 1000 =PQ030069.D 750 =PQ030070.D
 250 =PQ030072.D 50 =PQ030073.D

Compound			500	1000	750	250	50	Avg	%RSD	
6)	L1	AR-1016-4	3.331	2.698	3.322	3.417	2.894	3.132	E5	10.12
7)	L1	AR-1016-5	3.698	2.904	3.698	3.782	3.360	3.488	E5	10.46
8)	L2	AR-1221-1	1.607					1.607	E5	0.00
9)	L2	AR-1221-2	1.220					1.220	E5	0.00
10)	L2	AR-1221-3	3.730					3.730	E5	0.00
11)	L3	AR-1232-1	2.336					2.336	E5	0.00
12)	L3	AR-1232-2	1.889					1.889	E5	0.00
13)	L3	AR-1232-3	1.217					1.217	E5	0.00
14)	L3	AR-1232-4	1.326					1.326	E5	0.00
15)	L3	AR-1232-5	1.580					1.580	E5	0.00
16)	L4	AR-1242-1	2.670					2.670	E5	0.00
17)	L4	AR-1242-2	5.514					5.514	E5	0.00
18)	L4	AR-1242-3	2.600					2.600	E5	0.00
19)	L4	AR-1242-4	2.654					2.654	E5	0.00
20)	L4	AR-1242-5	2.952					2.952	E5	0.00
21)	L5	AR-1248-1	3.258					3.258	E5	0.00
22)	L5	AR-1248-2	3.513					3.513	E5	0.00
23)	L5	AR-1248-3	4.354					4.354	E5	0.00
24)	L5	AR-1248-4	6.482					6.482	E5	0.00
25)	L5	AR-1248-5	4.490					4.490	E5	0.00
26)	L6	AR-1254-1	7.476					7.476	E5	0.00
27)	L6	AR-1254-2	6.390					6.390	E5	0.00
28)	L6	AR-1254-3	1.102					1.102	E6	0.00
29)	L6	AR-1254-4	7.333					7.333	E5	0.00
30)	L6	AR-1254-5	4.284					4.284	E5	0.00
31)	L7	AR-1260-1	7.819	6.061	7.862	7.934	7.487	7.433	E5	10.57
32)	L7	AR-1260-2	9.036	7.361	9.141	9.064	8.358	8.592	E5	8.81
33)	L7	AR-1260-3	7.353	6.025	7.397	7.374	6.513	6.932	E5	9.09
34)	L7	AR-1260-4	1.889	1.575	1.897	1.870	1.632	1.773	E6	8.80
35)	L7	AR-1260-5	1.348	1.148	1.362	1.348	1.226	1.286	E6	7.38
36)	L8	AR-1262-1	5.937					5.937	E5	0.00
37)	L8	AR-1262-2	7.154					7.154	E5	0.00
38)	L8	AR-1262-3	1.068					1.068	E6	0.00
39)	L8	AR-1262-4	9.358					9.358	E5	0.00
40)	L8	AR-1262-5	2.027					2.027	E6	0.00
41)	L9	AR-1268-1	2.008					2.008	E6	0.00
42)	L9	AR-1268-2	1.942					1.942	E6	0.00
43)	L9	AR-1268-3	1.582					1.582	E6	0.00
44)	L9	AR-1268-4	6.797					6.797	E5	0.00
45)	L9	AR-1268-5	5.144					5.144	E6	0.00

(#) = Out of Range