

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 28 01:59:54 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ030145.D 1600 =PQ030147.D 800 =PQ030146.D
 200 =PQ030144.D 100 =PQ030143.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.318	4.292	4.215	4.396	4.684	4.381	E7	4.14
2) SA	Decachlorobiphenyl	3.027	2.908	3.112	3.122	3.377	3.109	E7	5.54
3) L1	AR-1016-1	7.119	6.595	6.759	7.350	7.584	7.081	E5	5.77
4) L1	AR-1016-2	0.961	0.946	0.947	0.974	1.009	0.967	E6	2.70
5) L1	AR-1016-3	1.244	1.200	1.305	1.266	1.355	1.274	E6	4.66
6) L1	AR-1016-4	1.068	1.052	1.106	1.090	1.116	1.086	E6	2.44
7) L1	AR-1016-5	8.759	8.596	9.075	9.065	9.808	9.061	E5	5.13
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	1.895	1.880	1.947	1.987	2.169	1.976	E6	5.88
32) L7	AR-1260-2	2.183	2.192	2.276	2.271	2.430	2.270	E6	4.36
33) L7	AR-1260-3	1.578	1.598	1.615	1.625	1.763	1.636	E6	4.47
34) L7	AR-1260-4	3.501	3.721	3.641	3.584	3.836	3.657	E6	3.51
35) L7	AR-1260-5	2.244	2.304	2.328	2.294	2.463	2.327	E6	3.53
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ030145.D 1600 =PQ030147.D 800 =PQ030146.D
 200 =PQ030144.D 100 =PQ030143.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.274	1.310	1.285	1.306	1.323	1.300	E7	1.52
2) SA	Decachlorobiphenyl	1.576	1.578	1.549	1.653	1.798	1.631	E7	6.21
3) L1	AR-1016-1	2.207	2.035	2.115	2.345	2.375	2.216	E5	6.57
4) L1	AR-1016-2	3.381	3.210	3.289	3.602	3.805	3.457	E5	7.04
5) L1	AR-1016-3	4.823	4.654	4.723	5.184	5.423	4.961	E5	6.63

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 28 01:59:54 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ030145.D 1600 =PQ030147.D 800 =PQ030146.D
 200 =PQ030144.D 100 =PQ030143.D

Compound			400	1600	800	200	100	Avg	%RSD	
6)	L1	AR-1016-4	3.560	3.518	3.479	3.723	3.808	3.618	E5	3.91
7)	L1	AR-1016-5	2.790	2.672	2.694	2.946	3.077	2.836	E5	6.09
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	8.119	7.927	7.919	8.619	9.264	8.370	E5	6.88
32)	L7	AR-1260-2	1.008	0.995	0.991	1.064	1.138	1.039	E6	6.01
33)	L7	AR-1260-3	0.937	0.936	0.925	0.980	1.041	0.964	E6	4.98
34)	L7	AR-1260-4	1.942	1.977	1.935	2.000	2.121	1.995	E6	3.77
35)	L7	AR-1260-5	1.376	1.399	1.369	1.420	1.513	1.415	E6	4.11
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

(#) = Out of Range