

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 28 02:44:24 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ030162.D 1600 =PQ030164.D 800 =PQ030163.D
 200 =PQ030161.D 100 =PQ030160.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.875	4.777	4.964	5.133	6.155	5.181	E7	10.81
2) SA	Decachlorobiphenyl	3.504	3.434	3.454	3.687	4.390	3.694	E7	10.88
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	1.816	1.752	1.827	1.922	2.380	1.939	E6	13.08
27) L6	AR-1254-2	2.794	2.719	2.840	2.937	3.671	2.992	E6	12.95
28) L6	AR-1254-3	2.996	2.985	3.077	3.136	4.099	3.259	E6	14.54
29) L6	AR-1254-4	2.231	2.213	2.293	2.371	2.875	2.397	E6	11.46
30) L6	AR-1254-5	1.739	1.729	1.802	1.865	2.131	1.853	E6	8.89
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ030162.D 1600 =PQ030164.D 800 =PQ030163.D
 200 =PQ030161.D 100 =PQ030160.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.467	1.432	1.476	1.473	1.579	1.485	E7	3.72
2) SA	Decachlorobiphenyl	1.896	1.783	1.869	1.952	2.166	1.933	E7	7.43
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062718CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 28 02:44:24 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ030162.D 1600 =PQ030164.D 800 =PQ030163.D
 200 =PQ030161.D 100 =PQ030160.D

Compound			400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4							0.000	-1.00
7) L1	AR-1016-5							0.000	-1.00
8) L2	AR-1221-1							0.000	-1.00
9) L2	AR-1221-2							0.000	-1.00
10) L2	AR-1221-3							0.000	-1.00
11) L3	AR-1232-1							0.000	-1.00
12) L3	AR-1232-2							0.000	-1.00
13) L3	AR-1232-3							0.000	-1.00
14) L3	AR-1232-4							0.000	-1.00
15) L3	AR-1232-5							0.000	-1.00
16) L4	AR-1242-1							0.000	-1.00
17) L4	AR-1242-2							0.000	-1.00
18) L4	AR-1242-3							0.000	-1.00
19) L4	AR-1242-4							0.000	-1.00
20) L4	AR-1242-5							0.000	-1.00
21) L5	AR-1248-1							0.000	-1.00
22) L5	AR-1248-2							0.000	-1.00
23) L5	AR-1248-3							0.000	-1.00
24) L5	AR-1248-4							0.000	-1.00
25) L5	AR-1248-5							0.000	-1.00
26) L6	AR-1254-1		0.956	0.888	0.936	0.986	1.090	0.971 E6	7.73
27) L6	AR-1254-2		8.376	7.775	8.185	8.667	9.626	8.526 E5	8.15
28) L6	AR-1254-3		1.426	1.344	1.405	1.460	1.600	1.447 E6	6.59
29) L6	AR-1254-4		0.941	0.892	0.929	0.960	1.049	0.954 E6	6.15
30) L6	AR-1254-5		6.601	6.382	6.610	6.691	7.434	6.744 E5	5.97
31) L7	AR-1260-1							0.000	-1.00
32) L7	AR-1260-2							0.000	-1.00
33) L7	AR-1260-3							0.000	-1.00
34) L7	AR-1260-4							0.000	-1.00
35) L7	AR-1260-5							0.000	-1.00
36) L8	AR-1262-1							0.000	-1.00
37) L8	AR-1262-2							0.000	-1.00
38) L8	AR-1262-3							0.000	-1.00
39) L8	AR-1262-4							0.000	-1.00
40) L8	AR-1262-5							0.000	-1.00
41) L9	AR-1268-1							0.000	-1.00
42) L9	AR-1268-2							0.000	-1.00
43) L9	AR-1268-3							0.000	-1.00
44) L9	AR-1268-4							0.000	-1.00
45) L9	AR-1268-5							0.000	-1.00

 (#) = Out of Range