

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ063018\
 Data File : PQ030220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2018 09:26
 Operator : UA/SM
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254305

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 00:33:46 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062718CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 30 00:25:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.704	850.1E6	242.9E6	19.403	18.687
2) SA Decachlor...	10.117	8.675	1395.8E6	694.3E6	44.891	42.573
Target Compounds						
26) L6 AR-1254-1	6.445	5.574	645.6E6	339.7E6	332.904	349.773
27) L6 AR-1254-2	6.661	5.719	970.6E6	293.9E6	324.378	344.667
28) L6 AR-1254-3	7.024	6.120	1042.8E6	502.8E6	320.006	347.399
29) L6 AR-1254-4	7.307	6.345	765.0E6	337.4E6	319.206	353.633
30) L6 AR-1254-5	7.579	6.659	614.7E6	236.7E6	331.757	351.031

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ063018\
Data File : PQ030220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2018 09:26
Operator : UA/SM
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254305

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 00:33:46 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062718CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 30 00:25:03 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

