

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070218\
 Data File : PQ030279.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2018 18:27
 Operator : UA/SM
 Sample : PB110715BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PB110715BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 05:36:53 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 29 01:53:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.433	3.700	817.4E6	286.3E6	19.474	20.805
2) SA Decachlor...	10.113	8.672	453.5E6	277.7E6	12.836	13.603

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070218\
Data File : PQ030279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2018 18:27
Operator : UA/SM
Sample : PB110715BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PB110715BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 05:36:53 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 29 01:53:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.433 min
Delta R.T.: -0.004 min
Response: 817396213
Conc: 19.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PB110715BL




#1 Tetrachloro-m-xylene

R.T.: 3.700 min
Delta R.T.: -0.003 min
Response: 286325695
Conc: 20.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.113 min
Delta R.T.: -0.006 min
Response: 453511435
Conc: 12.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.672 min
Delta R.T.: -0.004 min
Response: 277728977
Conc: 13.60 ng/ml