

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071619\
 Data File : PQ041252.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jul 2019 22:09
 Operator : SM\AJ
 Sample : K3733-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 KY030LC051-SD-190709

Manual Integrations
 APPROVED

Ankita
 7/21/2019 11:43:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 02:41:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 13 02:14:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.713	4.769	98040442	72573560	9.652	11.129
2) SA Decachlor...	12.127	10.539	109.5E6	46575609	8.524	9.963
Target Compounds						
31) L7 AR-1260-1	8.922	7.935	12662975	10593590	23.943m	31.178m#
32) L7 AR-1260-2	9.194	8.138	14507772	11597326	19.356	26.079 #
33) L7 AR-1260-3	9.568	8.304	12687528	10540969	19.779m	28.266m#
34) L7 AR-1260-4	9.809	8.803	15731769	6304534	25.929m	20.964
35) L7 AR-1260-5	10.162	9.052	24761692	16529851	17.351	22.322 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071619\
Data File : PQ041252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2019 22:09
Operator : SM\AJ
Sample : K3733-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
KY030LC051-SD-190709

Manual Integrations
APPROVED

Ankita
7/21/2019 11:43:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 02:41:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071219.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 13 02:14:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

