

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080818.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 09 08:35:15 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ031036.D 1000 =PQ031034.D 750 =PQ031035.D
 250 =PQ031037.D 50 =PQ031038.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.469	1.553	1.741	1.665	1.475	1.581	E6	7.58
2) SA	Decachlorobiphenyl	1.777	1.769	1.579	1.755	1.797	1.735	E6	5.11
3) L1	AR-1016-1	2.516	2.528	2.913	2.836	2.481	2.655	E4	7.65
4) L1	AR-1016-2	3.840	3.956	4.423	4.529	3.719	4.093	E4	8.82
5) L1	AR-1016-3	5.616	5.879	6.484	6.538	6.062	6.116	E4	6.45
6) L1	AR-1016-4	5.128	5.280	5.715	6.283	5.305	5.542	E4	8.44
7) L1	AR-1016-5	4.686	5.004	5.240	5.281	4.654	4.973	E4	5.96
8) L2	AR-1221-1	1.633					1.633	E4	0.00
9) L2	AR-1221-2	1.225					1.225	E4	0.00
10) L2	AR-1221-3	4.131					4.131	E4	0.00
11) L3	AR-1232-1	3.645					3.645	E4	0.00
12) L3	AR-1232-2	1.920					1.920	E4	0.00
13) L3	AR-1232-3	2.766					2.766	E4	0.00
14) L3	AR-1232-4	2.569					2.569	E4	0.00
15) L3	AR-1232-5	1.876					1.876	E4	0.00
16) L4	AR-1242-1	3.571					3.571	E4	0.00
17) L4	AR-1242-2	5.118					5.118	E4	0.00
18) L4	AR-1242-3	5.118					5.118	E4	0.00
19) L4	AR-1242-4	4.077					4.077	E4	0.00
20) L4	AR-1242-5	4.486					4.486	E4	0.00
21) L5	AR-1248-1	6.627					6.627	E4	0.00
22) L5	AR-1248-2	5.552					5.552	E4	0.00
23) L5	AR-1248-3	6.007					6.007	E4	0.00
24) L5	AR-1248-4	7.697					7.697	E4	0.00
25) L5	AR-1248-5	8.097					8.097	E4	0.00
26) L6	AR-1254-1	9.302					9.302	E4	0.00
27) L6	AR-1254-2	1.474					1.474	E5	0.00
28) L6	AR-1254-3	9.210					9.210	E4	0.00
29) L6	AR-1254-4	1.228					1.228	E5	0.00
30) L6	AR-1254-5	1.563					1.563	E5	0.00
31) L7	AR-1260-1	1.048	1.035	1.095	1.154	1.271	1.121	E5	8.56
32) L7	AR-1260-2	1.278	1.268	1.324	1.396	1.604	1.374	E5	10.04
33) L7	AR-1260-3	1.579	1.539	1.596	1.721	1.859	1.659	E5	7.88
34) L7	AR-1260-4	2.265	2.292	2.252	2.393	2.470	2.334	E5	4.02
35) L7	AR-1260-5	1.384	1.376	1.328	1.420	1.538	1.409	E5	5.62
36) L8	AR-1262-1	9.646					9.646	E4	0.00
37) L8	AR-1262-2	1.160					1.160	E5	0.00
38) L8	AR-1262-3	1.142					1.142	E5	0.00
39) L8	AR-1262-4	1.490					1.490	E5	0.00
40) L8	AR-1262-5	2.842					2.842	E5	0.00
41) L9	AR-1268-1	2.909					2.909	E5	0.00
42) L9	AR-1268-2	2.568					2.568	E5	0.00
43) L9	AR-1268-3	2.134					2.134	E5	0.00
44) L9	AR-1268-4	8.753					8.753	E4	0.00
45) L9	AR-1268-5	6.274					6.274	E5	0.00

Signal #2 Calibration Files

500 =PQ031036.D 1000 =PQ031034.D 750 =PQ031035.D
 250 =PQ031037.D 50 =PQ031038.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.408	1.503	1.698	1.596	1.413	1.524	E6	8.15
2) SA	Decachlorobiphenyl	1.953	1.909	1.760	1.930	2.058	1.922	E6	5.58
3) L1	AR-1016-1	3.611	3.704	4.213	4.292	3.957	3.955	E4	7.60
4) L1	AR-1016-2	5.640	5.936	6.566	6.694	6.202	6.208	E4	7.03
5) L1	AR-1016-3	8.179	8.502	9.418	9.431	8.270	8.760	E4	7.05

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080818.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 09 08:35:15 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ031036.D 1000 =PQ031034.D 750 =PQ031035.D
 250 =PQ031037.D 50 =PQ031038.D

	Compound	500	1000	750	250	50	Avg	%RSD
6)	L1 AR-1016-4	5.912	5.532	6.130	6.960	4.099	5.726 E4	18.33
7)	L1 AR-1016-5	3.412	3.356	3.813	4.031	4.197	3.762 E4	9.87
8)	L2 AR-1221-1	1.644					1.644 E4	0.00
9)	L2 AR-1221-2	1.262					1.262 E4	0.00
10)	L2 AR-1221-3	4.245					4.245 E4	0.00
11)	L3 AR-1232-1	3.602					3.602 E4	0.00
12)	L3 AR-1232-2	4.110					4.110 E4	0.00
13)	L3 AR-1232-3	2.955					2.955 E4	0.00
14)	L3 AR-1232-4	1.975					1.975 E4	0.00
15)	L3 AR-1232-5	2.410					2.410 E4	0.00
16)	L4 AR-1242-1	3.657					3.657 E4	0.00
17)	L4 AR-1242-2	7.689					7.689 E4	0.00
18)	L4 AR-1242-3	4.016					4.016 E4	0.00
19)	L4 AR-1242-4	3.242					3.242 E4	0.00
20)	L4 AR-1242-5	4.629					4.629 E4	0.00
21)	L5 AR-1248-1	5.629					5.629 E4	0.00
22)	L5 AR-1248-2	5.736					5.736 E4	0.00
23)	L5 AR-1248-3	7.182					7.182 E4	0.00
24)	L5 AR-1248-4	1.090					1.090 E5	0.00
25)	L5 AR-1248-5	7.245					7.245 E4	0.00
26)	L6 AR-1254-1	1.339					1.339 E5	0.00
27)	L6 AR-1254-2	1.180					1.180 E5	0.00
28)	L6 AR-1254-3	2.095					2.095 E5	0.00
29)	L6 AR-1254-4	1.332					1.332 E5	0.00
30)	L6 AR-1254-5	1.941					1.941 E5	0.00
31)	L7 AR-1260-1	1.181	1.143	1.235	1.297	1.421	1.255 E5	8.72
32)	L7 AR-1260-2	1.477	1.426	1.519	1.598	1.659	1.536 E5	6.07
33)	L7 AR-1260-3	1.208	1.161	1.210	1.267	1.432	1.255 E5	8.42
34)	L7 AR-1260-4	2.749	2.770	2.716	2.823	2.876	2.787 E5	2.28
35)	L7 AR-1260-5	1.973	1.930	1.892	2.037	2.053	1.977 E5	3.47
36)	L8 AR-1262-1	1.064					1.064 E5	0.00
37)	L8 AR-1262-2	1.274					1.274 E5	0.00
38)	L8 AR-1262-3	1.778					1.778 E5	0.00
39)	L8 AR-1262-4	1.603					1.603 E5	0.00
40)	L8 AR-1262-5	3.345					3.345 E5	0.00
41)	L9 AR-1268-1	3.382					3.382 E5	0.00
42)	L9 AR-1268-2	3.090					3.090 E5	0.00
43)	L9 AR-1268-3	2.525					2.525 E5	0.00
44)	L9 AR-1268-4	1.011					1.011 E5	0.00
45)	L9 AR-1268-5	7.162					7.162 E5	0.00

(#) = Out of Range