

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 10 05:08:33 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ031058.D 1600 =PQ031060.D 800 =PQ031059.D
 200 =PQ031057.D 100 =PQ031056.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.826	1.877	1.679	1.649	1.756	1.758	E6	5.47
2) SA	Decachlorobiphenyl	1.880	1.823	1.962	2.216	2.255	2.027	E6	9.71
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1	2.144	2.102	2.027	2.063	2.303	2.128	E4	5.03
17) L4	AR-1242-2	3.566	3.600	3.430	3.676	4.238	3.702	E4	8.45
18) L4	AR-1242-3	1.713	1.760	1.692	1.920	2.270	1.871	E4	12.84
19) L4	AR-1242-4	1.261	1.301	1.250	1.529	1.507	1.370	E4	9.99
20) L4	AR-1242-5	4.551	4.566	4.344	4.848	5.059	4.674	E4	5.99
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ031058.D 1600 =PQ031060.D 800 =PQ031059.D
 200 =PQ031057.D 100 =PQ031056.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.938	1.941	1.747	1.607	1.826	1.812	E6	7.74
2) SA	Decachlorobiphenyl	2.128	2.161	2.273	2.503	2.521	2.317	E6	8.02
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 10 05:08:33 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ031058.D 1600 =PQ031060.D 800 =PQ031059.D
 200 =PQ031057.D 100 =PQ031056.D

	Compound	400	1600	800	200	100	Avg		%RSD

6)	L1 AR-1016-4						0.000		-1.00
7)	L1 AR-1016-5						0.000		-1.00
8)	L2 AR-1221-1						0.000		-1.00
9)	L2 AR-1221-2						0.000		-1.00
10)	L2 AR-1221-3						0.000		-1.00
11)	L3 AR-1232-1						0.000		-1.00
12)	L3 AR-1232-2						0.000		-1.00
13)	L3 AR-1232-3						0.000		-1.00
14)	L3 AR-1232-4						0.000		-1.00
15)	L3 AR-1232-5						0.000		-1.00
16)	L4 AR-1242-1	2.206	2.166	2.046	2.030	2.255	2.140	E4	4.64
17)	L4 AR-1242-2	4.288	4.503	4.131	3.934	4.043	4.180	E4	5.32
18)	L4 AR-1242-3	3.387	3.363	3.184	3.235	3.217	3.277	E4	2.79
19)	L4 AR-1242-4	3.853	3.944	3.850	4.145	4.847	4.128	E4	10.16
20)	L4 AR-1242-5	3.281	3.534	3.367	3.406	3.957	3.509	E4	7.59
21)	L5 AR-1248-1						0.000		-1.00
22)	L5 AR-1248-2						0.000		-1.00
23)	L5 AR-1248-3						0.000		-1.00
24)	L5 AR-1248-4						0.000		-1.00
25)	L5 AR-1248-5						0.000		-1.00
26)	L6 AR-1254-1						0.000		-1.00
27)	L6 AR-1254-2						0.000		-1.00
28)	L6 AR-1254-3						0.000		-1.00
29)	L6 AR-1254-4						0.000		-1.00
30)	L6 AR-1254-5						0.000		-1.00
31)	L7 AR-1260-1						0.000		-1.00
32)	L7 AR-1260-2						0.000		-1.00
33)	L7 AR-1260-3						0.000		-1.00
34)	L7 AR-1260-4						0.000		-1.00
35)	L7 AR-1260-5						0.000		-1.00
36)	L8 AR-1262-1						0.000		-1.00
37)	L8 AR-1262-2						0.000		-1.00
38)	L8 AR-1262-3						0.000		-1.00
39)	L8 AR-1262-4						0.000		-1.00
40)	L8 AR-1262-5						0.000		-1.00
41)	L9 AR-1268-1						0.000		-1.00
42)	L9 AR-1268-2						0.000		-1.00
43)	L9 AR-1268-3						0.000		-1.00
44)	L9 AR-1268-4						0.000		-1.00
45)	L9 AR-1268-5						0.000		-1.00

 (#) = Out of Range