

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 10 04:41:55 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ031068.D 1600 =PQ031070.D 800 =PQ031069.D
 200 =PQ031067.D 100 =PQ031066.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.838	1.880	1.931	1.971	1.783	1.881	E6	3.95
2) SA	Decachlorobiphenyl	2.104	2.001	1.946	2.156	2.368	2.115	E6	7.76
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	1.585	1.451	1.518	1.686	1.754	1.599	E5	7.68
27) L6	AR-1254-2	0.968	0.904	0.934	1.031	1.083	0.984	E5	7.41
28) L6	AR-1254-3	1.751	1.669	1.700	1.857	1.963	1.788	E5	6.77
29) L6	AR-1254-4	1.282	1.217	1.241	1.387	1.408	1.307	E5	6.61
30) L6	AR-1254-5	0.987	0.950	0.946	1.033	1.098	1.003	E5	6.36
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ031068.D 1600 =PQ031070.D 800 =PQ031069.D
 200 =PQ031067.D 100 =PQ031066.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.974	2.058	2.139	2.139	1.808	2.024	E6	6.85
2) SA	Decachlorobiphenyl	2.500	2.369	2.371	2.513	2.768	2.504	E6	6.50
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 10 04:41:55 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ031068.D 1600 =PQ031070.D 800 =PQ031069.D
 200 =PQ031067.D 100 =PQ031066.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.422	1.255	1.391	1.486	1.531	1.417 E5	7.46
27) L6	AR-1254-2	1.198	1.100	1.194	1.228	1.258	1.196 E5	4.97
28) L6	AR-1254-3	1.774	1.660	1.754	1.789	1.913	1.778 E5	5.11
29) L6	AR-1254-4	1.075	1.006	1.035	1.100	1.143	1.072 E5	5.02
30) L6	AR-1254-5	2.222	2.054	2.110	2.294	2.350	2.206 E5	5.60
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range