

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082218\
 Data File : PQ031326.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2018 14:49
 Operator : SM\SJ
 Sample : J4271-11
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 HR-06-082018-B

Manual Integrations
 APPROVED

Sohil
 8/23/2018 10:43:49 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 04:20:09 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 22 08:49:07 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.599	3.881	32531053	34083484	15.634m	19.042m
2) SA Decachlor...	10.454	8.965	20768591	18915891	10.999	9.951

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082218\
Data File : PQ031326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2018 14:49
Operator : SM\SJ
Sample : J4271-11
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
HR-06-082018-B


Manual Integrations
APPROVED

Sohil
8/23/2018 10:43:49 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 04:20:09 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082018.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 22 08:49:07 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.599 min
Delta R.T.: 0.000 min
Response: 32531053
Conc: 15.63 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
HR-06-082018-B


Manual Integrations
APPROVED

Sohil
8/23/2018 10:43:49 AM




#1 Tetrachloro-m-xylene

R.T.: 3.881 min
Delta R.T.: 0.000 min
Response: 34083484
Conc: 19.04 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.454 min
Delta R.T.: 0.000 min
Response: 20768591
Conc: 11.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.965 min
Delta R.T.: -0.002 min
Response: 18915891
Conc: 9.95 ng/ml