

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0091918.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 20 02:02:28 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0049137.D 1000 =P0049135.D 750 =P0049136.D
 250 =P0049138.D 50 =P0049139.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.943	1.800	1.835	1.636	2.070	1.857	E5	8.74
2) SA	Decachlorobiphenyl	2.709	2.478	2.574	2.685	3.298	2.749	E5	11.66
3) L1	AR-1016-1	4.893	4.433	4.580	4.308	5.362	4.715	E3	8.96
4) L1	AR-1016-2	5.027	4.530	4.733	4.415	5.483	4.838	E3	8.87
5) L1	AR-1016-3	7.694	6.841	7.110	6.888	8.851	7.477	E3	11.23
6) L1	AR-1016-4	6.719	5.995	6.228	5.874	7.687	6.501	E3	11.35
7) L1	AR-1016-5	5.667	5.185	5.337	4.919	5.799	5.381	E3	6.64
8) L2	AR-1221-1	1.988					1.988	E3	0.00
9) L2	AR-1221-2	1.442					1.442	E3	0.00
10) L2	AR-1221-3	5.043					5.043	E3	0.00
11) L3	AR-1232-1	3.904					3.904	E2	0.00
12) L3	AR-1232-2	2.013					2.013	E3	0.00
13) L3	AR-1232-3	5.340					5.340	E2	0.00
14) L3	AR-1232-4	4.811					4.811	E2	0.00
15) L3	AR-1232-5	3.237					3.237	E2	0.00
16) L4	AR-1242-1	2.024					2.024	E3	0.00
17) L4	AR-1242-2	3.712					3.712	E3	0.00
18) L4	AR-1242-3	1.780					1.780	E3	0.00
19) L4	AR-1242-4	1.365					1.365	E3	0.00
20) L4	AR-1242-5	4.749					4.749	E3	0.00
21) L5	AR-1248-1	6.045					6.045	E3	0.00
22) L5	AR-1248-2	6.423					6.423	E3	0.00
23) L5	AR-1248-3	8.418					8.418	E3	0.00
24) L5	AR-1248-4	7.897					7.897	E3	0.00
25) L5	AR-1248-5	5.196					5.196	E3	0.00
26) L6	AR-1254-1	1.480					1.480	E4	0.00
27) L6	AR-1254-2	9.173					9.173	E3	0.00
28) L6	AR-1254-3	1.634					1.634	E4	0.00
29) L6	AR-1254-4	1.283					1.283	E4	0.00
30) L6	AR-1254-5	9.244					9.244	E3	0.00
31) L7	AR-1260-1	1.152	1.015	1.054	1.041	1.330	1.119	E4	11.53
32) L7	AR-1260-2	1.383	1.234	1.274	1.225	1.592	1.342	E4	11.44
33) L7	AR-1260-3	1.577	1.419	1.473	1.421	1.803	1.539	E4	10.46
34) L7	AR-1260-4	1.141	1.034	1.066	1.009	1.222	1.094	E4	7.94
35) L7	AR-1260-5	2.425	2.254	2.293	2.227	2.787	2.397	E4	9.63
36) L8	AR-1262-1	6.531					6.531	E3	0.00
37) L8	AR-1262-2	5.375					5.375	E3	0.00
38) L8	AR-1262-3	5.746					5.746	E3	0.00
39) L8	AR-1262-4	1.372					1.372	E4	0.00
40) L8	AR-1262-5	1.042					1.042	E4	0.00
41) L9	AR-1268-1	5.774					5.774	E3	0.00
42) L9	AR-1268-2	7.631					7.631	E3	0.00
43) L9	AR-1268-3	3.388					3.388	E4	0.00
44) L9	AR-1268-4	2.863					2.863	E4	0.00
45) L9	AR-1268-5	9.559					9.559	E4	0.00

Signal #2 Calibration Files

500 =P0049137.D 1000 =P0049135.D 750 =P0049136.D
 250 =P0049138.D 50 =P0049139.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.084	1.942	2.011	1.871	2.425	2.067	E5	10.42
2) SA	Decachlorobiphenyl	1.979	1.783	1.873	1.900	2.701	2.047	E5	18.17
3) L1	AR-1016-1	5.718	5.111	5.453	5.382	6.951	5.723	E3	12.57
4) L1	AR-1016-2	0.858	0.776	0.817	0.806	1.092	0.870	E4	14.69
5) L1	AR-1016-3	1.159	1.049	1.111	1.074	1.377	1.154	E4	11.37

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0091918.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 20 02:02:28 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0049137.D 1000 =P0049135.D 750 =P0049136.D
 250 =P0049138.D 50 =P0049139.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	8.253	7.421	7.850	7.537	9.092	8.031	E3	8.41
7)	L1 AR-1016-5	7.005	6.275	6.608	6.749	7.973	6.922	E3	9.31
8)	L2 AR-1221-1	2.430					2.430	E3	0.00
9)	L2 AR-1221-2	1.801					1.801	E3	0.00
10)	L2 AR-1221-3	5.586					5.586	E3	0.00
11)	L3 AR-1232-1	2.341					2.341	E3	0.00
12)	L3 AR-1232-2	8.568					8.568	E2	0.00
13)	L3 AR-1232-3	2.175					2.175	E3	0.00
14)	L3 AR-1232-4	2.372					2.372	E3	0.00
15)	L3 AR-1232-5	2.490					2.490	E2	0.00
16)	L4 AR-1242-1	2.609					2.609	E3	0.00
17)	L4 AR-1242-2	5.400					5.400	E3	0.00
18)	L4 AR-1242-3	1.794					1.794	E3	0.00
19)	L4 AR-1242-4	1.587					1.587	E3	0.00
20)	L4 AR-1242-5	4.172					4.172	E3	0.00
21)	L5 AR-1248-1	6.976					6.976	E3	0.00
22)	L5 AR-1248-2	1.300					1.300	E4	0.00
23)	L5 AR-1248-3	9.252					9.252	E3	0.00
24)	L5 AR-1248-4	8.116					8.116	E3	0.00
25)	L5 AR-1248-5	6.131					6.131	E3	0.00
26)	L6 AR-1254-1	1.227					1.227	E4	0.00
27)	L6 AR-1254-2	1.070					1.070	E4	0.00
28)	L6 AR-1254-3	1.376					1.376	E4	0.00
29)	L6 AR-1254-4	1.024					1.024	E4	0.00
30)	L6 AR-1254-5	1.939					1.939	E4	0.00
31)	L7 AR-1260-1	1.423	1.278	1.331	1.314	1.843	1.438	E4	16.17
32)	L7 AR-1260-2	1.743	1.568	1.623	1.604	2.326	1.773	E4	17.83
33)	L7 AR-1260-3	1.646	1.488	1.530	1.490	2.077	1.646	E4	15.13
34)	L7 AR-1260-4	1.323	1.186	1.230	1.217	1.706	1.332	E4	16.13
35)	L7 AR-1260-5	3.134	2.817	2.903	2.830	3.978	3.132	E4	15.63
36)	L8 AR-1262-1	7.138					7.138	E3	0.00
37)	L8 AR-1262-2	7.337					7.337	E3	0.00
38)	L8 AR-1262-3	7.769					7.769	E3	0.00
39)	L8 AR-1262-4	1.218					1.218	E4	0.00
40)	L8 AR-1262-5	1.017					1.017	E4	0.00
41)	L9 AR-1268-1	7.322					7.322	E3	0.00
42)	L9 AR-1268-2	3.320					3.320	E4	0.00
43)	L9 AR-1268-3	2.740					2.740	E4	0.00
44)	L9 AR-1268-4	7.773					7.773	E3	0.00
45)	L9 AR-1268-5	7.414					7.414	E4	0.00

(#) = Out of Range