

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092618CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 27 05:35:41 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032240.D 1600 =PQ032242.D 800 =PQ032241.D
 200 =PQ032239.D 100 =PQ032238.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.660	1.891	1.394	1.627	1.541	1.623	E6	11.22
2) SA	Decachlorobiphenyl	2.437	2.346	2.155	2.557	2.597	2.418	E6	7.35
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	0.952	1.016	0.811	0.988	1.036	0.960	E5	9.31
27) L6	AR-1254-2	1.555	1.641	1.321	1.619	1.636	1.555	E5	8.68
28) L6	AR-1254-3	1.776	1.881	1.505	1.837	1.815	1.763	E5	8.46
29) L6	AR-1254-4	1.332	1.377	1.118	1.365	1.366	1.311	E5	8.36
30) L6	AR-1254-5	1.518	1.543	1.283	1.540	1.554	1.488	E5	7.74
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ032240.D 1600 =PQ032242.D 800 =PQ032241.D
 200 =PQ032239.D 100 =PQ032238.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.293	1.532	1.106	1.188	1.126	1.249	E6	13.95
2) SA	Decachlorobiphenyl	2.168	2.170	1.961	2.240	2.314	2.171	E6	6.06
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092618CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 27 05:35:41 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032240.D 1600 =PQ032242.D 800 =PQ032241.D
 200 =PQ032239.D 100 =PQ032238.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.149	1.249	0.991	1.161	1.147	1.140 E5	8.15
27) L6	AR-1254-2	1.019	1.091	0.875	1.049	1.092	1.025 E5	8.70
28) L6	AR-1254-3	1.864	1.989	1.604	1.898	1.886	1.848 E5	7.84
29) L6	AR-1254-4	1.215	1.268	1.035	1.240	1.205	1.193 E5	7.67
30) L6	AR-1254-5	1.723	1.770	1.478	1.710	1.732	1.683 E5	6.92
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range