

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092618CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 27 08:19:03 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032245.D 1600 =PQ032247.D 800 =PQ032246.D
 200 =PQ032244.D 100 =PQ032243.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.400	1.284	1.330	1.093	1.573	1.336 E6	13.07	
2) SA	Decachlorobiphenyl	2.150	2.056	2.050	2.145	2.473	2.175 E6	7.97	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	
6) L1	AR-1016-4						0.000	-1.00	
7) L1	AR-1016-5						0.000	-1.00	
8) L2	AR-1221-1						0.000	-1.00	
9) L2	AR-1221-2						0.000	-1.00	
10) L2	AR-1221-3						0.000	-1.00	
11) L3	AR-1232-1						0.000	-1.00	
12) L3	AR-1232-2						0.000	-1.00	
13) L3	AR-1232-3						0.000	-1.00	
14) L3	AR-1232-4						0.000	-1.00	
15) L3	AR-1232-5						0.000	-1.00	
16) L4	AR-1242-1						0.000	-1.00	
17) L4	AR-1242-2						0.000	-1.00	
18) L4	AR-1242-3						0.000	-1.00	
19) L4	AR-1242-4						0.000	-1.00	
20) L4	AR-1242-5						0.000	-1.00	
21) L5	AR-1248-1						0.000	-1.00	
22) L5	AR-1248-2						0.000	-1.00	
23) L5	AR-1248-3						0.000	-1.00	
24) L5	AR-1248-4						0.000	-1.00	
25) L5	AR-1248-5						0.000	-1.00	
26) L6	AR-1254-1						0.000	-1.00	
27) L6	AR-1254-2						0.000	-1.00	
28) L6	AR-1254-3						0.000	-1.00	
29) L6	AR-1254-4						0.000	-1.00	
30) L6	AR-1254-5						0.000	-1.00	
31) L7	AR-1260-1						0.000	-1.00	
32) L7	AR-1260-2						0.000	-1.00	
33) L7	AR-1260-3						0.000	-1.00	
34) L7	AR-1260-4						0.000	-1.00	
35) L7	AR-1260-5						0.000	-1.00	
36) L8	AR-1262-1	1.613	1.502	1.540	1.453	1.871	1.596 E5	10.31	
37) L8	AR-1262-2	3.019	2.928	2.925	2.701	3.279	2.970 E5	7.02	
38) L8	AR-1262-3	1.997	1.883	1.896	1.897	2.213	1.977 E5	7.06	
39) L8	AR-1262-4	1.474	1.380	1.395	1.406	1.651	1.461 E5	7.67	
40) L8	AR-1262-5	1.045	0.985	0.992	1.007	1.172	1.040 E5	7.42	
41) L9	AR-1268-1						0.000	-1.00	
42) L9	AR-1268-2						0.000	-1.00	
43) L9	AR-1268-3						0.000	-1.00	
44) L9	AR-1268-4						0.000	-1.00	
45) L9	AR-1268-5						0.000	-1.00	

Signal #2 Calibration Files

400 =PQ032245.D 1600 =PQ032247.D 800 =PQ032246.D
 200 =PQ032244.D 100 =PQ032243.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.092	1.005	1.051	0.822	1.184	1.031 E6	13.03	
2) SA	Decachlorobiphenyl	1.930	1.873	1.866	1.902	2.189	1.952 E6	6.92	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092618CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 27 08:19:03 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032245.D 1600 =PQ032247.D 800 =PQ032246.D
 200 =PQ032244.D 100 =PQ032243.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	1.620	1.505	1.545	1.399	1.754	1.564 E5	8.48
37) L8	AR-1262-2	3.114	3.003	3.030	2.777	3.278	3.041 E5	5.99
38) L8	AR-1262-3	1.166	1.092	1.118	1.098	1.275	1.150 E5	6.57
39) L8	AR-1262-4	2.225	2.153	2.161	2.035	2.404	2.195 E5	6.17
40) L8	AR-1262-5	1.017	0.969	0.976	0.972	1.121	1.011 E5	6.37
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range