

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092618CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 27 08:04:25 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032250.D 1600 =PQ032252.D 800 =PQ032251.D
 200 =PQ032249.D 100 =PQ032248.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.454	1.607	1.451	1.464	1.592	1.513	E6	5.20
2) SA	Decachlorobiphenyl	4.585	4.644	4.489	4.921	5.119	4.752	E6	5.48
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	4.379	4.474	4.411	4.640	4.974	4.575	E5	5.34
42) L9	AR-1268-2	3.941	4.071	3.973	4.201	4.507	4.139	E5	5.55
43) L9	AR-1268-3	3.377	3.486	3.404	3.579	3.852	3.539	E5	5.41
44) L9	AR-1268-4	1.467	1.431	1.429	1.608	1.697	1.527	E5	7.89
45) L9	AR-1268-5	1.104	1.138	1.104	1.148	1.162	1.131	E6	2.32

Signal #2 Calibration Files

400 =PQ032250.D 1600 =PQ032252.D 800 =PQ032251.D
 200 =PQ032249.D 100 =PQ032248.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.127	1.264	1.148	1.109	1.253	1.180	E6	6.18
2) SA	Decachlorobiphenyl	4.101	4.292	4.118	4.372	4.504	4.277	E6	4.00
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ092618CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 27 08:04:25 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032250.D 1600 =PQ032252.D 800 =PQ032251.D
 200 =PQ032249.D 100 =PQ032248.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	4.320	4.443	4.403	4.585	4.854	4.521 E5	4.64
42) L9	AR-1268-2	3.867	4.060	3.960	4.084	4.306	4.055 E5	4.05
43) L9	AR-1268-3	3.194	3.392	3.303	3.353	3.584	3.365 E5	4.26
44) L9	AR-1268-4	1.410	1.422	1.394	1.554	1.596	1.475 E5	6.30
45) L9	AR-1268-5	1.032	1.070	1.048	1.082	1.074	1.061 E6	1.96

(#) = Out of Range