

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ100918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 10 05:28:12 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032775.D 1600 =PQ032777.D 800 =PQ032776.D
 200 =PQ032774.D 100 =PQ032773.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.953	2.169	1.912	2.198	2.501	2.147	E6	10.95
2) SA	Decachlorobiphenyl	2.335	1.834	1.815	2.387	2.526	2.180	E6	15.21
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	4.573	4.560	4.083	5.635	5.951	4.960	E4	15.99
22) L5	AR-1248-2	6.867	6.304	5.948	8.308	9.349	7.355	E4	19.48
23) L5	AR-1248-3	0.780	0.724	0.675	0.894	1.013	0.817	E5	16.69
24) L5	AR-1248-4	0.895	0.832	0.771	1.036	1.150	0.937	E5	16.48
25) L5	AR-1248-5	0.856	0.786	0.738	0.987	1.073	0.888	E5	15.72
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ032775.D 1600 =PQ032777.D 800 =PQ032776.D
 200 =PQ032774.D 100 =PQ032773.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.474	1.680	1.469	1.718	1.953	1.659	E6	12.07
2) SA	Decachlorobiphenyl	1.963	1.624	1.555	1.949	2.058	1.830	E6	12.29
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ100918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Oct 10 05:28:12 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ032775.D 1600 =PQ032777.D 800 =PQ032776.D
 200 =PQ032774.D 100 =PQ032773.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	3.826	3.782	3.374	4.469	5.039	4.098 E4	16.01
22) L5	AR-1248-2	5.080	4.863	4.505	6.134	6.888	5.494 E4	17.97
23) L5	AR-1248-3	5.186	5.022	4.638	6.275	6.880	5.600 E4	16.76
24) L5	AR-1248-4	6.508	6.295	5.721	7.578	8.383	6.897 E4	15.50
25) L5	AR-1248-5	6.686	6.306	5.730	7.687	8.155	6.913 E4	14.39
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range