Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101518\
Data File : PQ033182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2018 02:26
Operator : SM\SJ
Sample : MDL-AR1262-W-4
Misc :
ALS Vial : 142 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
MDL-AR1262-W-4

Manual Integrations APPROVED
sohil
10/19/2018 5:37:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 05:32:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 10 08:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.572	3.851	51493688	39493743	25.101	25.392m
2)	SA Decachlor...	10.405	8.903	78666314	68908477	41.616	42.865
	Target Compounds						
36)	L8 AR-1262-1	7.970	6.979	3480867	4004754	25.825m	31.162
37)	L8 AR-1262-2	8.532	7.480	6245438	7452578	25.257	31.113
38)	L8 AR-1262-3	8.851	7.763	3863812	3260554	23.359m	35.004 #
39)	L8 AR-1262-4	8.948	7.828	2404456	5092213	35.281m	29.250
40)	L8 AR-1262-5	9.630	8.334	2339460	2207531	26.740m	27.356m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101518\
Data File : PQ033182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2018 02:26
Operator : SM\SJ
Sample : MDL-AR1262-W-4
Misc :
ALS Vial : 142 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
MDL-AR1262-W-4

Manual Integrations APPROVED
sohil
10/19/2018 5:37:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 05:32:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 10 08:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm