

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101719\
 Data File : PQ044472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2019 13:53
 Operator : HP\AJ
 Sample : K5308-05
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 X2J23

Manual Integrations
 APPROVED

Ankita
 10/21/2019 10:18:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 08:45:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 10 02:55:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.330	4.350	88708628	97992984	22.818	21.953
2)	SA Decachlor...	11.624	9.801	179.6E6	214.7E6	33.531m	35.899
Target Compounds							
26)	L6 AR-1254-1	7.572	6.519	85036880	166.4E6	631.974	570.426
27)	L6 AR-1254-2	7.802	6.677	130.5E6	146.3E6	610.898	560.291
28)	L6 AR-1254-3	8.188	7.107	138.0E6	247.5E6	560.740	566.582
29)	L6 AR-1254-4	8.485	7.347	106.2E6	146.8E6	529.743	546.502
30)	L6 AR-1254-5	8.920	7.781	118.5E6	225.3E6	495.006	563.460

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101719\
Data File : PQ044472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2019 13:53
Operator : HP\AJ
Sample : K5308-05
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
X2J23

Manual Integrations
APPROVED

Ankita
10/21/2019 10:18:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 08:45:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ100819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 02:55:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

