

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ110618.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 09 08:12:53 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ034151.D 1000 =PQ034149.D 750 =PQ034150.D
 250 =PQ034152.D 50 =PQ034153.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.883	3.216	2.942	2.635	2.980	2.931	E6	7.11
2) SA	Decachlorobiphenyl	2.464	2.459	2.476	2.486	2.951	2.567	E6	8.37
3) L1	AR-1016-1	0.879	1.058	0.915	0.877	1.011	0.948	E5	8.68
4) L1	AR-1016-2	1.265	1.582	1.359	1.289	1.564	1.412	E5	10.72
5) L1	AR-1016-3	7.499	9.369	7.957	7.826	9.835	8.497	E4	12.19
6) L1	AR-1016-4	6.253	7.768	6.564	6.194	7.165	6.789	E4	9.86
7) L1	AR-1016-5	6.332	7.791	6.546	6.465	8.136	7.054	E4	11.95
8) L2	AR-1221-1	2.864					2.864	E4	0.00
9) L2	AR-1221-2	2.131					2.131	E4	0.00
10) L2	AR-1221-3	6.794					6.794	E4	0.00
11) L3	AR-1232-1	5.154					5.154	E4	0.00
12) L3	AR-1232-2	2.809					2.809	E4	0.00
13) L3	AR-1232-3	5.683					5.683	E4	0.00
14) L3	AR-1232-4	2.792					2.792	E4	0.00
15) L3	AR-1232-5	2.007					2.007	E4	0.00
16) L4	AR-1242-1	7.860					7.860	E4	0.00
17) L4	AR-1242-2	1.169					1.169	E5	0.00
18) L4	AR-1242-3	7.074					7.074	E4	0.00
19) L4	AR-1242-4	5.741					5.741	E4	0.00
20) L4	AR-1242-5	6.464					6.464	E4	0.00
21) L5	AR-1248-1	6.491					6.491	E4	0.00
22) L5	AR-1248-2	9.322					9.322	E4	0.00
23) L5	AR-1248-3	1.067					1.067	E5	0.00
24) L5	AR-1248-4	1.215					1.215	E5	0.00
25) L5	AR-1248-5	1.169					1.169	E5	0.00
26) L6	AR-1254-1	1.130					1.130	E5	0.00
27) L6	AR-1254-2	1.793					1.793	E5	0.00
28) L6	AR-1254-3	1.944					1.944	E5	0.00
29) L6	AR-1254-4	1.431					1.431	E5	0.00
30) L6	AR-1254-5	1.574					1.574	E5	0.00
31) L7	AR-1260-1	1.244	1.493	1.274	1.298	1.585	1.379	E5	10.97
32) L7	AR-1260-2	1.540	1.885	1.578	1.607	1.995	1.721	E5	11.91
33) L7	AR-1260-3	0.960	1.123	0.969	0.973	1.068	1.019	E5	7.17
34) L7	AR-1260-4	1.174	1.399	1.215	1.260	1.448	1.299	E5	9.13
35) L7	AR-1260-5	2.307	2.672	2.384	2.375	2.731	2.494	E5	7.75
36) L8	AR-1262-1	1.842					1.842	E5	0.00
37) L8	AR-1262-2	3.401					3.401	E5	0.00
38) L8	AR-1262-3	2.221					2.221	E5	0.00
39) L8	AR-1262-4	8.430					8.430	E4	0.00
40) L8	AR-1262-5	1.137					1.137	E5	0.00
41) L9	AR-1268-1	4.407					4.407	E5	0.00
42) L9	AR-1268-2	3.971					3.971	E5	0.00
43) L9	AR-1268-3	3.411					3.411	E5	0.00
44) L9	AR-1268-4	1.369					1.369	E5	0.00
45) L9	AR-1268-5	1.140					1.140	E6	0.00

Signal #2 Calibration Files

500 =PQ034151.D 1000 =PQ034149.D 750 =PQ034150.D
 250 =PQ034152.D 50 =PQ034153.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.869	2.072	1.911	1.720	1.971	1.909	E6	6.80
2) SA	Decachlorobiphenyl	1.867	1.855	1.891	1.905	2.308	1.965	E6	9.82
3) L1	AR-1016-1	6.448	7.704	6.791	6.447	7.928	7.064	E4	9.98
4) L1	AR-1016-2	0.931	1.117	0.974	0.911	1.107	1.008	E5	9.70
5) L1	AR-1016-3	4.851	5.885	5.082	4.821	5.039	5.135	E4	8.46

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ110618.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 09 08:12:53 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ034151.D 1000 =PQ034149.D 750 =PQ034150.D
 250 =PQ034152.D 50 =PQ034153.D

Compound			500	1000	750	250	50	Avg		%RSD
6)	L1	AR-1016-4	3.818	4.622	3.997	3.936	4.204	4.115	E4	7.68
7)	L1	AR-1016-5	5.046	6.161	5.275	5.140	5.044	5.333	E4	8.86
8)	L2	AR-1221-1	2.096					2.096	E4	0.00
9)	L2	AR-1221-2	1.547					1.547	E4	0.00
10)	L2	AR-1221-3	5.104					5.104	E4	0.00
11)	L3	AR-1232-1	3.784					3.784	E4	0.00
12)	L3	AR-1232-2	3.974					3.974	E4	0.00
13)	L3	AR-1232-3	2.057					2.057	E4	0.00
14)	L3	AR-1232-4	1.853					1.853	E4	0.00
15)	L3	AR-1232-5	2.021					2.021	E4	0.00
16)	L4	AR-1242-1	5.527					5.527	E4	0.00
17)	L4	AR-1242-2	7.996					7.996	E4	0.00
18)	L4	AR-1242-3	4.128					4.128	E4	0.00
19)	L4	AR-1242-4	4.024					4.024	E4	0.00
20)	L4	AR-1242-5	5.221					5.221	E4	0.00
21)	L5	AR-1248-1	4.637					4.637	E4	0.00
22)	L5	AR-1248-2	6.274					6.274	E4	0.00
23)	L5	AR-1248-3	6.468					6.468	E4	0.00
24)	L5	AR-1248-4	7.969					7.969	E4	0.00
25)	L5	AR-1248-5	8.044					8.044	E4	0.00
26)	L6	AR-1254-1	1.156					1.156	E5	0.00
27)	L6	AR-1254-2	1.001					1.001	E5	0.00
28)	L6	AR-1254-3	1.728					1.728	E5	0.00
29)	L6	AR-1254-4	1.090					1.090	E5	0.00
30)	L6	AR-1254-5	1.517					1.517	E5	0.00
31)	L7	AR-1260-1	1.064	1.274	1.084	1.069	1.263	1.151	E5	9.36
32)	L7	AR-1260-2	1.299	1.553	1.340	1.322	1.585	1.420	E5	9.69
33)	L7	AR-1260-3	1.213	1.463	1.265	1.236	1.409	1.317	E5	8.47
34)	L7	AR-1260-4	8.449	9.691	8.697	8.637	9.896	9.074	E4	7.35
35)	L7	AR-1260-5	2.173	2.459	2.266	2.141	2.580	2.324	E5	8.15
36)	L8	AR-1262-1	1.502					1.502	E5	0.00
37)	L8	AR-1262-2	2.882					2.882	E5	0.00
38)	L8	AR-1262-3	1.068					1.068	E5	0.00
39)	L8	AR-1262-4	2.048					2.048	E5	0.00
40)	L8	AR-1262-5	9.115					9.115	E4	0.00
41)	L9	AR-1268-1	3.559					3.559	E5	0.00
42)	L9	AR-1268-2	3.195					3.195	E5	0.00
43)	L9	AR-1268-3	2.628					2.628	E5	0.00
44)	L9	AR-1268-4	1.112					1.112	E5	0.00
45)	L9	AR-1268-5	8.652					8.652	E5	0.00

(#) = Out of Range