

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ112918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 30 01:05:34 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ035065.D 1600 =PQ035067.D 800 =PQ035066.D
 200 =PQ035064.D 100 =PQ035063.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.118	4.398	3.830	3.667	3.820	3.767 E6	12.14
2) SA	Decachlorobiphenyl	3.905	3.006	3.822	4.094	4.366	3.839 E6	13.29
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	3.035	3.979	3.640	3.677	4.164	3.699 E4	11.62
9) L2	AR-1221-2	2.092	2.770	2.571	2.478	2.515	2.485 E4	9.93
10) L2	AR-1221-3	7.738	9.800	9.034	9.256	9.675	9.100 E4	9.04
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

400 =PQ035065.D 1600 =PQ035067.D 800 =PQ035066.D
 200 =PQ035064.D 100 =PQ035063.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.527	2.304	1.938	1.905	1.931	1.921 E6	14.32
2) SA	Decachlorobiphenyl	2.363	1.946	2.361	2.539	2.684	2.379 E6	11.63
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ112918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 30 01:05:34 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ035065.D 1600 =PQ035067.D 800 =PQ035066.D
 200 =PQ035064.D 100 =PQ035063.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	1.774	2.539	2.090	2.461	2.543	2.281 E4	14.88
9) L2	AR-1221-2	1.274	1.734	1.495	1.651	1.688	1.568 E4	11.96
10) L2	AR-1221-3	4.725	6.095	5.302	5.702	6.078	5.580 E4	10.36
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range